



CANADA

SD-WA.268

DEPARTMENT OF TRADE AND COMMERCE
STANDARDS DIVISION

OTTAWA, April 13th, 1956.

APPROVAL LISTING

Equal Arm Beam Scale Manufactured by
W & T Avery Co. Ltd., Birmingham, England.

Under the provisions of the Weights and Measures Act, Chapter 292, R.S.C. 1952, and Regulations thereunder (P.C. 6894), the apparatus specified and illustrated herein has been listed as an approved device and may be used in Canada in accordance with the conditions applicable.

Apparatus Listed: Beam over pan type equal arm counter scale, Model 2029 in capacities of 4 oz., 8 oz., 1 lb., 2 lbs., 7 lbs., and 10 lbs.

Application: Weighing commodities in trade.

Conditions: As prescribed in P.C. 6894.

Description: This device is a fine weighing equal arm counter balance with the pans suspended by stirrups from the load pivots. The beam also is suspended from the scale standard by a stirrup and the beam balance indicator is located in the stirrup slot when the scale is in balance. Pivots are hardened steel fitted in slots and bearings are V type agate. The base is cast iron and the scale is brass.

Testing: The standard tests for an equal arm balance shall apply.

C. S. Phillips,
Assistant Director (W&M),
Standards Division.

R. W. MacLean,
Director,
Standards Division.

REFERENCE:

A-391

AVERY EQUAL ARM COUNTER BALANCE

MODEL 2029

