



SD-WA.299

CANADA

DEPARTMENT OF TRADE AND COMMERCE  
STANDARDS DIVISION

OTTAWA, May 2, 1957

## APPROVAL LISTING

Bench type dial scale - manufactured by  
W. & T. Avery Co. Ltd., Birmingham, England.

Under the provisions of the Weights and Measures Act, Chapter 292, R.S.C. 1952, and Regulations thereunder (P.C. 6894), the apparatus specified and illustrated herein has been listed as an approved device and may be used in Canada in accordance with the conditions applicable.


**Apparatus Listed:** Bench type dial scale with taring and printing attachment, model 3303DFF, capacity 55 lbs. x 2 oz.


**Application:** Shipping, receiving, warehouse, and industrial use.

**Conditions:** As prescribed in P.C. 6894.

**Description:** This device is equipped with the same type of leverage system and registering head as model 3303CFB (SD-WA.237) except that: (1) the tare beam has been removed and replaced by a taring device operating on the registering head resistance spring. (2) a manually operated ticket printing device has been added and is so installed that the indicator must be at rest before printing can take place. A red dot on the control arm moves into alignment with an indicating line on a plastic window to warn the operator when the printing handle and control button may be operated. The printing ring prints in two sense intervals. The taring device control knob is located on the left side of the dial housing and when turned increases and decreases tension on the head resistance spring.

**Testings:** The Standard tests for a platform dial scale equipped with a printer and taring device shall apply.

  
C. S. Phillips,  
Assistant Director (W&M),  
Standards Division.

  
R. W. MacLean,  
Director,  
Standards Division.

**Reference:** A-452

SD-WA-299

Avery Bench Scale with Printer  
Model 3303 D.F.F.

Ticket printer

