



CANADA

DEPARTMENT OF TRADE AND COMMERCE
STANDARDS DIVISION

OTTAWA, June 28th, 1957

Approval Listing

Computing Counter scale - manufactured by
Bizerba - Waagen - Verkauf, G.M.B.H.,
Balingen (Wurt) Postfach 107, Germany.

Under the provisions of the Weights and Measures Act, Chapter 292, R.S.C. 1952, and Regulations thereunder (P.C.6894), the apparatus specified and illustrated herein has been listed as an approved device and may be used in Canada in accordance with the conditions applicable.

Apparatus Listed: Automatic indicating computing scale model U.S.18.


Rating of Apparatus: Weighing capacity 24 lbs., computing capacity 2 lbs.


Application: Weighing merchandise in retail trade.

Conditions: As prescribed in P.C.6894.

Description: This device is an automatic indicating scale, computing to 2 lbs. and weighing to 24 lbs. The dial capacity normally is 2 lbs. but is increased by increments of 2 lbs. by a hand operated lever on the side of the scale which moves a poise weight on a channel beam section of the main lever and locks it in the desired position. The chart has three windows at the weight reading line in which the appropriate pound numerals appear. The movement of the pound numerals is synchronized with the movement of the poise. The chart price range is from 10¢ to 1.25¢ and chart indications have 1¢ increments. Weight indications appear on the customers side. The main lever is of the unequal arm type. The scale is equipped with an adjustable dash pot and the whole mechanism is installed on a cast iron base and housed in a polished aluminum casing.

Testing: The standard tests for a computing scale shall apply.


C. S. PHILLIPS,
Assistant Director (W&M),
Standards Division.


R. W. MacLEAN,
Director,
Standards Division.

Reference: A-451

HEXAGON COUNTER SCALE
MODEL P. S. 12

