



CANADA

DEPARTMENT OF TRADE AND COMMERCE
STANDARDS DIVISION

OTTAWA,.....June 28th, 1957.....

APPROVAL LISTING

Equal arm counter scale - manufactured by
Bizerba - Waagen - Verkauf G.M.B.H
Balingen (Wurt) Postfach 107, Germany.

Under the provisions of the Weights and Measures Act, Chapter 292,
R.S.C. 1952, and Regulations thereunder (P.C.6894), the apparatus specified and
illustrated herein has been listed as an approved device and may be used in
Canada in accordance with the conditions applicable.


Apparatus Listed: Equal arm over and under indicating counter scale model RS,
capacity 6 lbs.


Application: Weighing commodities in retail trade or prepackaging.

Conditions: As prescribed in P.C.6894. NOTE: The over and under chart may
be graduated but weight values may not be marked thereon.

Description: This device is an equal arm scale with an end tower supporting
the over and under indicating mechanism. The load and weight
bridges are stabilized by a satisfactory linkage and a shot cup
is provided on each bridge. Oscillations are controlled by a
magnetic damping device located in the base of the end tower and
acting on an extension of the indicator shaft. The device is
installed on an aluminum alloy base, is encased in a polished
aluminum housing and is supported on three adjustable levelling
feet.

Testing: The standard tests for an equal arm counter scale shall apply.


C. S. Phillips,
Assistant Director (W&M),
Standards Division.


R. W. MacLean,
Director,
Standards Division.

Reference: A-445

Beierke Equal Arm Scale
Model 25

