

TRADE AND COMMERCE
CANADA

STANDARDS DIVISION

OTTAWA, February 18, 1959.

APPROVAL LISTING

SALTER SUSPENDED DIAL SCALE

Under the provisions of the Weights and Measures Act, Chapter 292, R.S.C. 1952, and Regulations thereunder (P.C. 6894), the apparatus specified and illustrated herein has been listed as an approved device and may be used in Canada in accordance with the conditions applicable.


Apparatus Listed: Suspended spring double faced dial scale Model 235T, capacity 30 lbs. x 1 oz., manufactured by George Salter and Company Limited, West Bromwich, England.


Application: Weighing commodities in general trade.

Conditions: As prescribed in P.C. 6894.

Description: This device is identical to scales listed on SD-WA. 170 in design, except that the dial reading line has been increased in diameter from 8 5/8 inches to 12 3/8 inches with relatively greater spacing of graduations.

Testing: The standard tests for a suspended dial scale shall apply.


C.S. Phillips,
Assistant Director (W&M),
Standards Division


R.W. MacLean,
Director,
Standards Division.

Reference: A-268A

)

)

