

TRADE AND COMMERCE  
CANADA

SD-WA. 370

## STANDARDS DIVISION

OTTAWA, February 24, 1959.

## APPROVAL LISTING

## AVERY - SUSPENDED DIAL SCALE

Under the provisions of the Weights and Measures Act, Chapter 292, R.S.C. 1952, and Regulations thereunder (P.C. 6894), the apparatus specified and illustrated herein has been listed as an approved device and may be used in Canada in accordance with the conditions applicable.


**Apparatus Listed:** Suspended dial scale model 1958-1-1CM with a capacity of 300 lbs., manufactured by W.&T. Avery Limited, Birmingham, England.


**Application:** Weighing wholesale quantities of goods in general trade and industry.

**Conditions:** As prescribed in P.C. 6894.

**Description:** This device is a spring and lever suspended dial scale. The registering head is basically the same as that installed on model 3205 C.F.M. (SD-WA. 298) except that two levers are added within the head casing to reduce the direct load on the head mechanism. The suspended load receiving element may be platform or scoop type.

**Testing:** The standard tests for a suspended dial scale shall apply.

  
C.S. Phillips,  
Assistant Director (W&M),  
Standards Division.

  
R.W. MacLean,  
Director,  
Standards Division.

**Reference:** A-511

=====

\_\_\_\_\_


=====

\_\_\_\_\_

(


(

✓



TYPE 1000 (SHOWING  
PRESSURE HEAD)

30-44-30



TYPE 1000 (SHOWING  
PRESSURE HEAD)

=====

-----

=====

(

(

4