



TRADE AND COMMERCE  
CANADA

## STANDARDS DIVISION

OTTAWA, July 16, 1959.

## APPROVAL LISTING

## SAUTER - PLATFORM DIAL BENCH SCALE

Under the provisions of the Weights and Measures Act, Chapter 292, R.S.C. 1952 and Regulations thereunder (P.C. 6894), the apparatus specified and illustrated herein has been listed as an approved device and may be used in Canada in accordance with the conditions applicable.

Apparatus Listed: Bench type platform dial scales including the following models:

B200V, B200W and B200X capacity 20-pounds by 1/4 ounce,  
B201V, B201W and B201X capacity 40-pounds by 1/2 ounce,  
B202V, B202W and B202X capacity 60-pounds by 1/2 ounce, manufactured by August Sauter K.G., Ebingen, West Germany. Branch Office: August Sauter of New York Incorporated, 866 Willis Avenue, Albertson, New York, U.S.A.

Application: General trade use, shipping and warehouse weighing.

Conditions: As prescribed in P.C. 6894.

NOTE: The listed models may be supplied with metric capacities comparative to the avoirdupois capacities listed.

Description: This device is a bench type automatic registering dial scale which requires five revolutions of the indicator to register capacity load. The head mechanism is similar to the scales listed on SD-WA.378. The underdeck leverage system consists of two first class A type levers located parallel to the sides and a transfer lever. The affixes V, W and X refer to different heights of the dial supporting column.

Testing: The standard tests for a platform dial scale shall apply.

*C.S. Phillips*  
C.S. Phillips,  
Assistant Director (W&M),  
Standards Division.

*R.W. MacLean*  
R.W. MacLean,  
Director,  
Standards Division

Reference: A-529

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

(



(


4

.

UNDER DECK LEVERAGE SYSTEM

3 2014





(

(

5



TRADE AND COMMERCE  
CANADA

SD-WA. 380

## STANDARDS DIVISION

OTTAWA.....July 23, 1959.....

## APPROVAL LISTING

## DETECTO - PLATFORM DIAL SCALE

Under the provisions of the Weights and Measures Act, Chapter 292, R.S.C. 1952, and Regulations thereunder (P.C. 6894), the apparatus specified and illustrated herein have been listed as approved devices and may be used in Canada in accordance with the conditions applicable.

Apparatus Listed: Platform dial scale with spring resistance, Models:


1105 D6V and 1106 DK with capacities of 200-pounds x 4-ounces and 300-pounds x 8-ounces, manufactured by Detecto Scales Incorporated, 540 Park Avenue, Brooklyn 5, New York, U.S.A.


Application: General trade and industrial use.

Conditions: As prescribed in P.C. 6894.

Description: These devices are platform dial scales with spring resistance and leverage system of the same construction as scales listed on SD-WA. 373 and only differ in capacity and chart markings.

Testing: The standard tests for a platform dial scale shall apply.

  
G.S. Phillips,  
Assistant Director (W&M),  
Standards Division.

  
R.W. MacLean,  
Director,  
Standards Division.

Reference: A-506

MODEL 1105 D6V

