

RECORD OF APPROVAL ACTION

Approval Listing Number - S.WA - T 100 (Temporary)

(Issued in advance of preparation of draft Approval Listing only in case of Interim Approval).

Company Requesting Approval - Accurate and Automatic Scale Co. Ltd.
2040 Ellesmere Road, Unit 15
Scarborough, Ontario M1H 3A8

Type of Device Single load cell platform scale.

<u>Model No.</u>	<u>Rating or Capacity</u>		<u>Platform Size</u>
BS-46-5	5 lb	2 kg	6" x 6"
BS-46-10	10 lb	5 kg	6"x8" or 8" x 10"
BS-46-30	30 lb	15 kg	12" x 16"
BS-46-50	50 lb	25 kg	14" x 18"
BS-46-100	100 lb	50 kg	18" x 20"

Application: The above models are stainless steel, platform scales to be equipped with the Accutronic 1000 or 2001 Digital Indicator. Load sensing is by means of a single Transducer Inc. series 46 load cell.

Special Conditions: The device shall be approved only on a temporary basis while being evaluated for full type approval in the Standards Laboratory.
No device shall be sold for use in trade without written disclosure to the purchaser, prior to the sale, that the device has received only temporary approval.

Date of Application for Approval - March 22, 1978

File No. G6900-A223

.....March 30, 1978...
Date of Filing

John Anderson
.....
Signature

NOTE: File Record of Approval Action with Approval Listing notices when Approval Listing number is assigned in advance of preparation of notice; file book of book by Central File No. when Approval Listing number is not assigned. Remove Record of Approval Action from Approval Listing Book when the company is notified of rejection, or when Special Approval Listing notice or draft of Approval Listing notice is filed with notices; replace draft of Approval Listing notice with printed notice when available.