



RECORD OF APPROVAL ACTION

Approval Listing Number -

T103

(Issued in advance of preparation of draft Approval Listing only  
in case of Interim Approval).

Ottawa, August 3, 1978

Company Requesting Approval - Colt Industries (Canada) Ltd.,  
91 Esna Park Drive,  
MARKHAM, Ontario.

Type of Device Fairbanks Built-In Electronic Scale

These scales are fully electronic utilizing 4 load cells supporting a platform  
consisting of a bridge assembly and check ~~up~~ plate decking. Load cells are  
column type under <sup>COMPRESSION</sup> tension. The scales are built-in and as such the overall  
dimensions will vary for different installations.

Conditions of Temporary Approval:

1. The Manufacturer must advise CSO of each new installation and provide all required documentation and drawings.
2. The scale must be used in conjunction with the Fairbanks H90-7302 indicator approved under SWA 913.
3. The approval to run for not greater than one year from date of issue.

Models presently installed:

SWA T 103

<u>MODEL</u>	<u>PLATFORM SIZE</u>	<u>CAPACITY</u>	<u>LOAD CELL</u>
23-2010	10 X 10 ft	30 tons	50 K pounds
23-2015 SP	8 X 5 ft	15 tons	50 K pounds

File No. G6922-F156

- 10

*John Armstrong*  
John Armstrong  
Chief

Weights and Measures Division  
Legal Metrology Branch

Note: File Record of Approval Action with Approval Listing notices when  
Approval Listing number is assigned in advance of preparation of notice; file at  
back of book by Central File No. when Approval Listing number is not assigned.

Remove Record of Approval Action from Approval Listing Book when the company is  
notified of rejection, or when Special Approval Listing notice or draft of  
Approval Listing notice is filed with notices; replace draft of Approval Listing  
notice with printed notice when available.