



Consumer and
Corporate Affairs

Consommation et
corporations

Standards

Normes

RECORD OF APPROVAL ACTION

Approval Listing Number - T104 Revised Dec. 13/78

(Issued in advance of preparation of draft Approval Listing only
in case of Interim Approval).

Ottawa, August 11, 1978

Company Requesting Approval - Data Terminal Systems,
124 Acton Road,
Maynard, Massachusetts,
U.S.A. 01754

Type of Device


DTS Model 925 load receiving element for use only with the Model 440 Electronic Cash Register, as all electronic functions for the load receiving element such as cell excitation, signal amplification, conversion and digital display are located in the ECR.

Conditions of Approval

1. All installations must be retrofitted with the B3 ROM or equivalent. When retrofitting is to be carried out on an installation, the Regional Weights and Measures Office must be informed.
2. This Approval is valid until June 1, 1979.

File No. G6922-087

JD


John Armstrong
Chief
Weights and Measures Division
Legal Metrology Branch

Note File Record of Approval Action with Approval Listing notices when Approval Listing number is assigned in advance of preparation of notice; file at back of book by Central File No. when Approval Listing number is not assigned.

Remove Record of Approval Action from Approval Listing Book when the company is notified of rejection, or when Special Approval Listing notice or draft of Approval Listing notice is filed with notices; replace draft of Approval Listing notice with printed notice when available.