



Consumer and  
Corporate Affairs

Consommation et  
corporations

Standards

Normes

RECORD OF APPROVAL ACTION

Approval Listing Number -

SWA-T111

(Issued in advance of preparation of draft Approval Listing only  
in case of Interim Approval).

Ottawa, December 1, 1978

Company Requesting Approval - Eastern Scale Magnatech Ltd.  
30 Carnforth Road, Toronto M4A 2K9

Type of Device NCI Model 5850-55 Dual Platform Counting Scale.  
This scale combines a small capacity ratio counting scale  
(100 lb.) with a high capacity (5000 lb.) platform scale.

Conditions of Approval:

- 1) This approval covers 3 devices only. Each device sold must be reported to the Chief, Weights and Measures by the manufacturer or distributor.
- 2) No device shall be sold for use in trade without written disclosure to the purchaser, prior to the sale, that the device has received only temporary approval.
- 3) The minimum allowable operating temperature is  $+10^{\circ}$  Celsius.
- 4) This approval to run for one year from date of issue.

File No. G6922-E82

*BE*

John Armstrong  
Chief

Weights and Measures Division  
Legal Metrology Branch

Note: File Record of Approval Action with Approval Listing notices when Approval Listing number is assigned in advance of preparation of notice; file at back of book by Central File No. when Approval Listing number is not assigned.

Remove Record of Approval Action from Approval Listing Book when the company is notified of rejection, or when Special Approval Listing notice or draft of Approval Listing notice is filed with notices; replace draft of Approval Listing notice with printed notice when available.



Consumer and  
Corporate Affairs

Consommation et  
corporations

Standards

Normes

Refer to  
X-180 Rev 1

RECORD OF APPROVAL ACTION

Approval Listing Number -

SWA-T111

(Issued in advance of preparation of draft Approval Listing only in case of Interim Approval).

Ottawa, December 1, 1978

Company Requesting Approval - Eastern Scale Magnatech Ltd.  
30 Carnforth Road, Toronto M4A 2K9

Type of Device NCI Model 5850-55 Dual Platform Counting Scale.  
This scale combines a small capacity ratio counting scale (100 lb.) with a high capacity (5000 lb.) platform scale.

Conditions of Approval:

1) This approval covers 3 devices only. Each device sold must be reported to the Chief, Weights and Measures by the manufacturer or distributor.

2) No device shall be sold for use in trade without written disclosure to the purchaser, prior to the sale, that the device has received only temporary approval.

3) The minimum allowable operating temperature is +10° Celsius.

4) This approval to run for one year from date of issue.

File No. G6922-E82

*se*

*John Armstrong*  
John Armstrong  
Chief

Weights and Measures Division  
Legal Metrology Branch

Note: File Record of Approval Action with Approval Listing notices when Approval Listing number is assigned in advance of preparation of notice; file at back of book by Central File No. when Approval Listing number is not assigned.

Remove Record of Approval Action from Approval Listing Book when the company is notified of rejection, or when Special Approval Listing notice or draft of Approval Listing notice is filed with notices; replace draft of Approval Listing notice with printed notice when available.