



Consumer and  
Corporate Affairs Canada

Consommation et  
Corporation Canada

Ottawa/Hull  
K1A 0C9

Consumer Standards  
Directorate

Direction générale  
des normes

RECORD OF APPROVAL ACTION

Approval listing Number -

SWA-T123

(Issued in advance of preparation of draft Approval Listing only  
in case of interim Approval).

Ottawa, May 16, 1979

Company Requesting Approval - Weigh Tronix, Inc  
1000 Armstrong Drive  
Fairmont, Minnesota 56031

Type of Device

Low profile and platform type weighbridges. These scales  
utilize four electronic Weighbars which are a particular design of  
electronic load cell.

Capacities of these scales varies from 500 to 50,000 pounds or  
metric equivalent. The Model number numeric designation refers to  
the size and capacity of the platform. Various combinations are  
possible. For example, the designation 6048-10 refers to a 60" X 48"  
deck with a rated capacity of 10000 pounds.


Model number prefixes are as follows:

- a) Low Profile - LP - Standard low profile with Std. WI102 indicator
  - LPF - Epoxy painted deck.
  - LPF - Special - stainless steel understructure.
  - LPFS - Totally stainless steel construction.
- b) Platform - DS - Standard dormant
  - DSF - Epoxy painted deck
  - DSFS - Totally stainless steel construction.

(See page 2)

File No. G6922-W75

*ee*

  
 .....  
 John Armstrong  
 Chief  
 Weights & Measures Division  
 Legal Metrology Branch

Conditions of Approval:

1. This approval covers 5 devices only. Each device sold must be reported to the Chief, Weights & Measures by the manufacturer. This report must include a brief description of the device, specifying capacities, sizes and options supplied. An outline drawing must be included.
2. No devices shall be sold for use in trade without written disclosure to the purchaser, prior to the sale, that the device has received only temporary approval.
3. This approval to run for one year from date of issue.
4. Applicant to submit test reports to the Chief, Weights and Measures, 3 months, and 8 months after the date of issue to this approval. Each test to consist of increasing/decreasing load test and section test generally per items 2 a) and 2 b) of the Weights & Measures Pre-Approval Test Requirement Guide.
5. Reports to be submitted by all W & M regional or district offices in whose regions an installation occurs. An initial verification report plus at least one control report to be submitted during the period of this approval by each relevant office.