



Consumer and
Corporate Affairs Canada

Consommation et
Corporation Canada

Ottawa/Hull
K1A 0C9

Consumer Standards
Directorate

Direction générale
des normes

RECORD OF APPROVAL ACTION

Approval listing Number -

SWA-T128

(Issued in advance of preparation of draft Approval Listing only
in case of interim Approval).

Ottawa, July 16, 1979

Company Requesting Approval - Donahue St. Felicien Inc.
St. Felicien, Quebec, G0W 2N0

Type of Device

Toledo Model 2154 Floor Scale with Toledo Model 8130 solid state digital readout is used to weigh paper rolls through a conveyor belt transport system. Bales of paper are transported by a conveyor belt at 75FPM and delivered to the scale platform. The scale is raised by either automatic or manual control - each bale is weighed and then the scale is lowered and the bales are then transported by a second conveyor belt system.

The capacity of the scale is 1250 lbs x 1.0 lbs.

File No. G6922-T170-10


.....
John Armstrong
Chief
Weights & Measures Division
Legal Metrology Branch

Conditions of Approval:

1. This approval covers one device only located at St. Felicien, Quebec.
2. This approval to run for one year from date of issue.
3. Applicant to submit test reports to the Chief, Weights and Measures, 3 months, and 8 months after the date of issue to this approval. Each test to consist of increasing/decreasing load test and section test generally per items 2 a) and 2 b) of the Weights & Measures Pre-Approval Test Requirement Guide.
4. An initial verification report plus at least one control report to be submitted during the period of this approval by the Quebec District Office.