



RECORD OF APPROVAL ACTION

Approval Listing Number -

SWA-T132

Revision 1

(Issued in advance of preparation of formal Approval Notice)

First Issued September 4, 1979.

Ottawa, January 18, 1980

Company Requesting Approval - DIGI-Digital Scales of Canada  
1 Hafis Road  
Toronto, Ontario M6M 2V6

Type of Device -

1. Type of Scale: Digital computing scale.
2. Manufactured by: Teraoka Seikosh Co. Ltd.  
5-13-12 Kugahara, Ohta-Ku, Tokyo, Japan
3. Description:

| <u>Model</u> | <u>Capacity/Increment</u>             | <u>Remarks</u>          |
|--------------|---------------------------------------|-------------------------|
| DS60-30      | 30 lb. or 15 kg./ .01 lb. or .005 kg. |                         |
| DS60-30D     | " " " " " " " " " "                   | Plus fractional pricing |
| DS60-15      | 15 lb. or 6 kg./ .005 lb. or .002 kg. |                         |
| DS60-15D     | " " " " " " " " " "                   | Plus fractional pricing |

4. Operational Features: Computes total price. Displays Net and Tare weights, unit and total price. Zero, Net, Prepack mode lites. Optional DP9100 printer may be used with scale.


Note: See Conditions of Approval on page 2.

File No. G6922-T192

.....  
 John Buchanan  
 Acting Chief  
 Weights & Measures Division  
 Legal Metrology Branch

Conditions of Approval:

1. This approval expires August 31, 1980.
2. The location and serial number of each device sold must be reported to the Chief, Weights and Measures by the manufacturer.
3. All devices covered by this approval shall be retrofitted before the expiry date of this approval so that the following areas of non-compliance have been corrected:
  - a) Cancellation of tare mode by depressing Tare key. This is contrary to section 14.5 of Ministerial Specification SGM1.
  - b) Excess range of auto zero maintenance. This is contrary to section 13.3 of Ministerial Specification SGM1.
4. These devices are exempt from section 15 of Ministerial Specification SGM1 for any weight up to 0.2 pounds placed on the platter.
5. This approval to be replaced by a special approval after all retrofitting has been completed. The manufacturer to advise the Chief, Weights and Measures when all retrofits have been completed.
5. This device is exempt from section 14.5 for cancellation of tare by removal of net load.

  
-----  
John Buchanan  
Acting Chief  
Weights and Measures Division  
Legal Metrology Branch