



Consumer and  
Corporate Affairs

Consommation et  
corporations

Standards

Normes

RECORD OF APPROVAL ACTION

Approval Listing Number -

SWA-X136

(Issued in advance of preparation of draft Approval Listing only  
in case of Interim Approval).

Ottawa, July 24, 1978

Company Requesting Approval -

Howe Richardson  
217 Brunswick Blvd  
Pointe Claire, Que. H9R 4R7


Type of Device Special Avery Model 5203 DLB Steel Mill Scale

located at Dosco Steel Ltd., Contrecoeur, Quebec.  
Consists of two standard type 5203 5'0" x 4' 0" self-  
contained weighbridge units of 10 tons each, positioned  
at 50'0" centre distances. Shock absorber assemblies  
are arranged on the platform, onto which the girders  
superstructure is fitted.

Unit No. 1 Serial No. BN64459-1 capacity 19990 lbs

Unit No. 2 Serial No. BN64459-2 Capacity 19990 lbs.

File No. G6922-A221-10V5

  
.....  
John Armstrong  
Chief  
Weights and Measures Division  
Legal Metrology Branch

Notes: File Record of Approval Action with Approval Listing notices when  
Approval Listing number is assigned in advance of preparation of notice; file at  
back of book by Central File No. when Approval Listing number is not assigned.

Remove Record of Approval Action from Approval Listing Book when the company is  
notified of rejection, or when Special Approval Listing notice or draft of  
Approval Listing notice is filed with notices; replace draft of Approval Listing  
notice with printed notice when available.