



Consumer and  
Corporate Affairs Canada

Consommation et  
Corporation Canada

Consumer Standards  
Directorate

Direction générale  
des normes

RECORD OF APPROVAL ACTION

Approval Listing Number -

SWA-T141

(Issued in advance of preparation of formal Approval Notice)

Ottawa, December 20, 1979

Company Requesting Approval - Mantle Industries Ltd.,  
8760 River Road, Delta, B.C.

Type of Device - This device is a full load cell paper roll scale using Transducer Load Cells B5102, 5K, a model 7200 digital weight indicator (refer to S.WA-1043), a remote readout and keyboard assembly with a continuous tape printer to print time, date, roll number and weight.

Model Number - PRS-1

Capacity and Readout Increments - 5000 kg x 1 kg

Conditions of Approval -

1. This approval covers 1 device only, the installation being at Prince George Pulp and Paper Ltd., Prince George, B.C.
2. No device shall be sold for use in trade without written disclosure to the purchaser, prior to the sale, that the device has received only temporary approval.
3. This approval to run for one year from date of issue.
4. An initial Weights and Measures test be performed in-situ.
5. Applicant to perform two tests (6 months and 12 months) after initial test and submit the results to the Chief, Weights and Measures. Each test to consist of increasing/decreasing load tests and section test, generally as per items 2 a) and 2 b) of the Weights and Measures Pre-Approval Test Requirement Guide.

File No. G6922-M36 *DL*

*John Armstrong*  
.....  
John Armstrong  
Chief  
Weights & Measures Division  
Legal Metrology Branch