



Consumer and
Corporate Affairs

Consommation et
corporations

Standards

Normes

**NOTICE OF APPROVAL
AVIS D'APPROBATION**

X144
S.WA-~~1139~~

Ottawa, November 17, 1978

GRIFFITHS - ELECTRONIC DIGITAL WEIGHT INDICATOR

MANUFACTURER: D.M. Griffiths Scales Ltd.
Burnaby, B.C.

DEVICE APPROVED: Digital Weight Indicator, Model LCR100
Readouts - gross weight, 5 digits
 - tare weight, 5 digits
 - net weight, 5 digits

Tare setting: keyboard, 10 pushbuttons 1 to 0
Output: binary coded decimal

CAPACITY: Maximum - 60,000 kg x 10 kg increments
 -125,000 lb. x 10 lb. increments

APPLICATION: For use in trade in conjunction with compatible
single or multiple load cell systems in approved weighing
installations.

SPECIAL CONDITIONS: This Special Approval Listing is applicable
to Griffiths LCR100 indicators presently installed at the following
locations only:

- | | | |
|-------------|--------------------------------|--------------------|
| One Unit | - Blackham Construction Ltd. | - Abbotsford, B.C. |
| Two Units | - City of Vancouver | - Vancouver, B.C. |
| Three Units | - Construction Aggregates Ltd. | - Vancouver |
| | | - North Vancouver |
| | | - Surrey, B.C. |
| One Unit | - Valley Gravel Sales Ltd. | - Aldergrove, B.C. |
| One Unit | - D.M. Griffiths Scales Ltd. | - Burnaby, B.C. |

TESTING: The standard tests for the applicable scale type shall apply.

REFERENCE: G6922-G314-33.

CONDITIONS OF APPROVAL: Approval is granted under the Weights and
Measures Act, S.C. 1970-71-72, Chapter 36, and the Weights and

CONDITIONS OF APPROVAL: (continued)

Measures Regulations P.C. 1974-1461 of June 27, 1974 for use in Canada under the general conditions of the said Regulations, and under any special conditions listed above.



John Armstrong
Chief, Weights and Measures
Legal Metrology Branch

