



Consumer and  
Corporate Affairs

Consommation et  
corporations

Standards

Normes

**NOTICE OF APPROVAL  
AVIS D'APPROBATION**

S.WA-X-151

Ottawa, February 26, 1980

Company Requesting Approval - Western Scale Co. Ltd.,  
Port Coquitlam, B.C.

Manufactured By - Western Scale Co. Ltd.,  
Port Coquitlam, B.C.

Type of Scale - Electronic Hopper Scale

Use of Scale - Bulk weighing of grain


Capacity - 11 000 kg

Model Number - HSTLC-4  
HS - Hopper Scale  
T - Tension  
LC - Load Cell  
4 - number of cells used

Device Description - This device is a full load cell automatic grain hopper scale, utilizing four (4) 10K Transducer Load Cells mounted in tension and interfaced to a Model DF100 Digital Weight Indicator (S.WA-1021).

Special Conditions: This approval covers the installation of ten (10) units. The manufacturer is to inform the chief of Legal Metrology and the Regional and/or District Office of Weights and Measures of the location of each unit.

.../2



Conditions of Approval: Approval is granted under the Weights and Measures Act, S.C. 1970-71-72, Chapter 36, and the Weights and Measures Regulations C.R.C. c., 1605 for use in Canada under the general conditions of the said Regulations, and under any conditions listed above.

Reference No.: G6922-W172-117

*J. D. Buchanan*  
.....  
John Buchanan  
Acting Chief  
Weights and Measures Division  
Legal Metrology Branch