



Consumer and
Corporate Affairs Canada

Consommation et
Corporation Canada

Consumer Standards
Directorate

Direction générale
des normes

RECORD OF APPROVAL ACTION

Approval Listing Number -

S.WA-T152

Rev. 1

Ottawa, December 1, 1981

Company Requesting Approval - Industrial Scale Co., Inc.
45 East Bend Ave., North
Hamilton, Ontario

Type of Device -

Industrial Platform scales, electronic type, utilizing four (4)
cantilever load cells, Type SB3, manufactured by Hottinger Baldwin
Measurement Inc.

MODEL LISTING - SERIES VP

Model No.	Capacity	Platform Size Feet	Load Cell Rating - Pounds
VP - 04.04.06 - 2 04.06.06 - 2	2000 lb x 0.5 lb or 1000 kg x 200 gm	4 x 4 4 x 6	1000
VP - 04.04.06 - 4 04.04.03 - 4 04.06.06 - 4 04.06.03 - 4 04.08.06 - 4	4000 lb x 1 lb or 2000 kg x 500 gm	4 x 4 4 x 4 4 x 6 4 x 6 4 x 8	2000
VP - 04.06.06 - 10 04.06.03 - 10 04.08.06 - 10 04.08.03 - 10 04.12.06 - 10 04.08.06 - 10 04.08.03 - 10 05.07.06 - 10	10,000 lb x 2 lb or 5000 kg x 1 kg	4 x 6 4 x 6 4 x 8 4 x 8 4 x 12 6 x 8 6 x 8 5 x 7	5000
VP - 06.08.03 - 20 VP - 06.08.06 - 20	20,000 lb x 5 lb or 10,000 kg x 2 kg	6 x 8 6 x 8	10,000

MODEL LISTING - SERIES VP

Model No.	Capacity	Platform Size Feet	Load Cell Rating - Pounds
04.06.08 - 40	40,000 lb x 5 lb	4 x 6	20,000
04.08.08 - 40	or 20,000 kg x 2 kg	4 x 8	
04.12.08 - 40		4 x 12	
04.24.08 - 40		4 x 24	
06.08.08 - 40		6 x 8	
06.12.08 - 40		6 x 12	
06.24.08 - 40		6 x 24	
06.06.12 - 100	100,000 lb x 10 lb	6 x 6	50,000
06.08.12 - 100	or 50,000 kg x 5 kg	6 x 8	
08.12.12 - 100		8 x 12	

Conditions of Approval:

1. This approval expires on March 31, 1982.
2. These scales may be used with approved compatible electronic indicators.
3. This approval covers 5 devices only. Each device sold must be reported to the Chief, Weights and Measures by the manufacturer. This report must include a brief description of the device, specifying capacities, sizes and options supplied. Two copies of the outline drawing must be included.
4. No devices shall be sold for use in trade without written disclosure to the purchaser, prior to the sale, that the device has received only temporary approval.
5. Reports to be submitted by all W & M regions or district offices in whose regions an installation occurs. An initial verification report plus at least one control report to be submitted during the period of this approval by each relevant office.
6. This approval also applies to installations conforming to the basic design above but having non-standard platform size and/or capacity specifications. In this case approval is subject to prior agreement between the manufacturer and the Chief, Weights and Measures.

File: G6922-J425

R. Bruce, Acting Chief
Chief, Weights and Measures
Legal Metrology Branch