



RECORD OF APPROVAL ACTION

Approval Listing Number -

S.WA-T156

Ottawa, October 2, 1980

Company Requesting Approval: Toledo Scale
Division of Reliance Electric Ltd.
5220 Creekbank Road, Mississauga

Type of Device: Self-contained electronic floor scales, with weight sensing by means of cantilever type load cells supporting the weighbridge, and readout by means of a digital indicator.

Model Number: 2295/xxxx. The latter portion of the model number serves to identify the indicator model.

Platform Size: 4' x 6', 6' x 8', 7' x 10'. Alternately, a scale may be custom built to suit a particular application within the maximum capacity rating.

Maximum Capacity: 40,000 lb. or 20,000 kg.

Conditions of Approval:


1. This approval expires on October 2, 1981.
2. These scales may be used with approved compatible electronic indicators.
3. This approval covers 10 devices only. Each device sold must be reported to the Chief, Weights and Measures by the manufacturer. This report must include a brief description of the device, specifying capacities, sizes and options supplied. Two copies of the outline drawing must be included.
4. No devices shall be sold for use in trade without written disclosure to the purchaser, prior to the sale, that the device has received only temporary approval.

Conditions of Approval: (Cont'd)

5. Reports to be submitted by all W&M regions or district offices in whose regions an installation occurs. An initial verification report plus at least one control report to be submitted during the period of this approval by each relevant office.

6. In the case where a scale is installed above the floor, it must be secured to the floor in a permanent manner.

File No.: G6922-T170-10



John Armstrong
Chief, Weights & Measures Division
Legal Metrology Branch