



N. H. Kelly

S.WA-T156 Rev. 1

Approval Listing Number

Ottawa, 14-10-81

RECORD OF APPROVAL ACTION
Formule d'enregistrement de l'approbation

Company:

Toledo Scales
Division of Reliance Electric Ltd
5220 Creebank Road
Mississauga, Ontario

Type of Device: Self-contained electronic floor scales, with weight sensing by means of cantilever type load cells supporting the weighbridge, and readout by means of a digital indicator.

Model Number: 2295/xxxx. The latter portion of the model number serves to identify the indicator model.

Platform Size: 4' x 6'; 6' x 8'; 7' x 10'. Alternately, a scale may be custom built to suit a particular application within the maximum capacity rating.

Maximum Capacity: 40 000 lb. or 20 000 kg.

Conditions of Approval:

- 1) This approval expires on January 31, 1982.
- 2) These scales may be used with approved compatible electronic indicators.

Société:

Toledo Scales
Division of Reliance Electric Ltd
5220 Creebank Road
Mississauga, Ontario

Appareil: Bascule de sol électronique autonome dans laquelle les dynamomètres de flexion sont à la fois éléments détecteurs de charge et éléments de soutien du châssis récepteur. L'affichage est réalisé au moyen d'un indicateur numérique.

Numéro de modèle: 2295/xxxx. La dernière tranche du numéro de modèle désigne le modèle de l'indicateur.

Dimensions de tablier: 4' x 6', 6' x 8', 7' x 10'. Les bascules peuvent aussi être conçues de façon à s'adapter à un usage particulier, compte tenu de la portée maximale spécifiée.

Portée maximale: 40 000 lb ou 20 000 kg.

Conditions d'approbation:

- 1) Le présent avis expire le 31 janvier 1982.
- 2) Les bascules qui font l'objet du présent avis peuvent être utilisées avec des indicateurs électroniques compatibles et approuvés.

Conditions of Approval: (Cont'd)

3) Each device sold must be reported to the Chief, Weights and Measures by the manufacturer. This report must include a brief description of the device, specifying capacities, sizes and options supplied. Two copies of the outline drawing must be included.

4) No device shall be sold for use in trade without written disclosure to the purchaser, prior to the sale, that the device has received only temporary approval.

5) Reports to be submitted by all W&M regions or district offices in whose regions an installation occurs. An initial verification report plus at least one control report to be submitted during the period of this approval by each relevant office.

6) In the case where a scale is installed above the floor, it must be secured to the floor in a permanent manner.

File Number: G6922-T170-10

Conditions d'approbation: (Suite)


3) Tout appareil vendu doit être signalé au chef de la Division des poids et mesures par le fabricant. Un tel rapport doit comprendre une courte description de l'appareil, sa portée, ses dimensions, les options offertes et deux exemplaires du schéma de l'appareil.

4) Aucun appareil destiné à être utilisé à des fins commerciales ne doit être vendu sans que l'acheteur n'ait d'abord été avisé par écrit que l'appareil n'a reçu qu'une approbation provisoire.

5) Toute installation doit être signalée dans un rapport provenant du bureau de Poids et mesures de la région ou du district où l'installation a eu lieu. Un rapport de vérification initiale, puis au moins un rapport de contrôle doivent être présentés par chacun des bureaux concernés au cours de la période visée par le présent avis.

6) Lorsqu'une bascule est installée au-dessus du sol, elle doit être fixée au sol de façon permanente.

Référence No.: G6922-T170-10


Robert Bruce, A/Chief
Weights and Measures Division
Legal Metrology

Robert Bruce, chef intérimaire
Division des poids et mesures
Métrologie légale