



S.WA-X157

Approval No. d'approbation

Ottawa June 2, 1980

NOTICE OF APPROVAL – AVIS D'APPROBATION

Manufacturer: Weightech Group
91 Esna Park Drive,
Markham, Ontario

Apparatus: Fully Electronic Tank Scales

Model Numbers:

E1076-4 (maximum gross capacity 60,000 lbs.)

E1077-4 (maximum gross capacity 90,000 lbs.)

Device Description: These devices are fully electronic tank scales,
Model E1076-4 uses four (4) 20k compression load cells (C93-50k-4009),
Model E1077-4 uses four (4) 50k compression load cells (C93-50k-4009),
these load cells are mounted in parallel link load cell mounts (S.WA-1008)
and interfaced to a Fairbanks Digital Indicator Model H90-7302 (S.WA-917).

The actual capacity of the device is determined by
the tank size.

Special Conditions: This approval applies only to model number E1076-4
installed at Centres Scientifiques Belisle Inc., 196 Chemin
des Patriotes, St. Mathias, Quebec, and model number E1077-4 installed
at Dominion Dairies Ltd., 235 Walmer Road, Toronto, Ontario.

Conditions of Approval: Approval is granted under the Weights and
Measures Act, S.C. 1970-71-72, Chapter 36, and the Weights and Measures
Regulations C.R.C.c., 1605, for use in Canada under the general conditions
of the said Regulations, and under any special conditions listed above.

File No. *JL* G6922-F158-11

.....*J. D. Buchanan*.....
John Buchanan
Acting Chief
Weights and Measures Division
Legal Metrology Branch