



Consumer and
Corporate Affairs

Consommation et
corporations

Legal Metrology

Métrologie Légale

S.WA-X160

Approval No. d'approbation

Ottawa July 22, 1980

NOTICE OF APPROVAL – AVIS D'APPROBATION

Company Requesting Approval: Fraser Valley Auctions
21801-56 Ave.,
Langley, B.C.

Scale Installer: Libra Scale Co., Ltd.
6984 Palm Ave.,
Burnaby, B.C.

Apparatus: Full electronic platform scale, 9 ft. x 10 ft., formerly being a Fairbanks lever scale, now converted to electronic by the installation of four load cells in conjunction with an electronic indicator.

Load Cells: Four(4) bending beam load cells (Weigh Bars), manufactured by Weigh-Tronix Inc., Model No. D15993-0056, Cap. 2500 lb/cell.

Indicator: Weigh-Tronix Model W1-1893. 5000 lb x 1 lb.

Capacity: 5000 pounds or metric equivalent.

Application: The weighing of livestock in trade.

Special Conditions: 1. This approval is valid only for the installation at Fraser Valley Auctions, Langley, B.C.

File *J.B.* G6922-W75

Conditions of Approval: Approval is granted under the Weights and Measures Act, S.C. 1970-71-72, Chapter 36, and the Weights and Measures Regulations (C.R.C., c. 1605) for use in Canada under the general conditions of the said Regulations, and under any special conditions listed above.

J. S. Buchanan

John Buchanan
Acting Chief
Weights and Measures
Legal Metrology Branch