



RECORD OF APPROVAL ACTION

Approval Listing Number

S.WA-T165

(Issued in advance of preparation of Formal Approval Notice)

OTTAWA, November 10, 1980

Manufacturer: Aurora Scale Mfg., Ltd.,
Aurora, Ontario

Company: Aurora Scale Mfg., Ltd.

Type of Device: Electronic Tank Scale

<u>MODEL NO.</u>	<u>CAPACITY</u>	<u>LOAD CELL MAKE, MODEL NUMBER AND CAPACITY</u>
ETA-15D3-200	200,000 lbs or metric equivalent	Revere CSPI-200 k

Application: The weighing of chemicals in trade.

Approved Accessories: This device may be interfaced to any approved compatible accessory.

Device Description: This device is a full electronic tank scale. The tank which is 15 ft. in diameter is supported by three (3) compression load cells. Weight indication is by means of an Analogic AN5322 (master indicator) and a remote indicator, a "Weighman" and a ticket printer are also incorporated.


To facilitate testing, a platform is suspended from brackets on the tank, this platform is capable of supporting twenty 10000 lb test weights.

Conditions of Approval: 1) This approval covers the device installed at Erco Chemical Ltd., Thunder Bay, Ontario.

2) No device shall be sold for use in trade without written disclosure to the purchaser, prior to the sale, that the device has received only temporary approval.

3) This approval to expire one year from date of issue.

FILE NO.: G6922-A364


John Armstrong, Chief
Weights and Measures Division
Legal Metrology Branch