



Consumer and
Corporate Affairs

Consommation et
corporations

Legal Metrology

Métrieologie Légale

S.WA-X167

Approval No. d'approbation

Ottawa October 1, 1980

NOTICE OF APPROVAL - AVIS D'APPROBATION

COMPANY REQUESTING APPROVAL: Donohue St. Felicien Inc.
St. Felicien, Quebec
GOW 2N0

APPARATUS: Automatic paper bale weighing system consisting of a Toledo model 2154 electronic platform scale with a Toledo model 8130 digital indicator. The scale is installed in a conveyor system.

A bale of paper is delivered by a conveyor to the scale platform, following which the scale platform is automatically elevated for the weighing operation, then lowered and the bale is removed by a second conveyor. The cycle is then repeated on a continuous basis.

SCALE MODEL NO.: Toledo 2154/8130.


SERIAL NO.: 8B47R-0

CAPACITY: 1250 pounds.

SPECIAL CONDITIONS: This approval is valid only for the above-listed system installed at Donohue St. Felicien Inc. St. Felicien, Quebec.

FILE: G6922-T170-10

CONDITIONS OF APPROVAL: Approval is granted under the Weights and Measures Act, S.C. 1970-71-72, Chapter 36, and the Weights and Measures Regulations (C. 1605) for use in Canada under the general conditions of the said Regulations and under any special conditions listed above.



John Armstrong
Chief, Weights and Measures
Legal Metrology Branch