



Consumer and
Corporate Affairs

Consommation et
corporations

Legal Metrology

Métrieologie Légale

P McEvedy

S.WA-X169

Approval No. d'approbation

Ottawa October 20, 1980

NOTICE OF APPROVAL – AVIS D'APPROBATION

COMPANY REQUESTING APPROVAL: Mantle Industries Ltd.
8760 River Road
Delta, B.C. V4C 3N3

TYPE OF DEVICE: Paper roll scale, full load cell type with digital weight indicator, remote readout and keyboard assembly and tape printer.

MODEL NO.: PRS-1

PLATFORM SIZE: 7 ft. x 8 ft.

CAPACITY: 5000 kg x 1 kg


LOAD CELLS: 4, cantilever type, 5K, Transducers Inc. B5102.

INDICATOR: Mantle/Finlayson Model 7200, Ref. S.WA-1043.

SPECIAL CONDITIONS: This approval is valid only for the above-listed installation located at Prince George Pulp and Paper Ltd., Prince George, B.C.

CONDITIONS OF APPROVAL: Approval is granted under the Weights and Measures Act, S.C. 1970-71-72, Chapter 36, and the Weights and Measures Regulations (C. 1605) for use in Canada under the general conditions of the said Regulations and under any special conditions listed above.

FILE: G6922-M365


John Armstrong
Chief, Weights and Measures
Legal Metrology Branch