



*Refer to
SWA-3196*

RECORD OF APPROVAL ACTION

Approval Listing Number

S.WA-T171

(Issued in advance of preparation of Formal Approval Notice)

OPTAWA, December 22, 1980

Company: Canadian Scale Company Limited
305 Horner Avenue
Toronto, Ontario

Manufacturer: Canadian Scale Company Ltd.

Type of Device: Railroad track scale.

<u>MODEL NUMBER</u>	<u>SECTIONAL CAPACITY</u>	<u>CAPACITY</u>
MRT-428-150	120 tons	300,000 lbs. or metric equivalent

Approved Accessories: This device may be used with any approved compatible accessory.


Device Description: This device is a four section mechanical lever live deck railroad track scale.

The levers are "A" shaped and the weighbridge is suspended by a double parallel link system. Indication is by means of an existing Fairbanks-Morse Weighbeam.

Conditions of Approval:

- 1) This approval covers the device at Alexander Centres Ltd., Sudbury, Ontario.
- 2) No device shall be sold for use in trade without written disclosure to the purchaser, prior to the sale, that the device has received only temporary approval.
- 3) This approval to expire one year from date of issue.

File No.: ^{SWA} G6922-C257



John Armstrong,
Chief
Weights and Measures Division
Legal Metrology Branch



Refer to
S.W.A-3196

S.W.A-T171 Rev. 1
Approval Listing Number

Ottawa, 1982/02/18

NOTICE OF TEMPORARY APPROVAL
Avis d'approbation temporaire

THIS NOTICE OF TEMPORARY APPROVAL SUPERSEDES RECORD OF APPROVAL ACTION S.W.A-T171 DATED December 22, 1980.

LE PRESENT AVIS D'APROBATION TEMPORAIRE REMPLACE LA CIRCULAIRE S.W.A-T171 EMISE LE 22 décembre 1980.

Company: Canadian Scale Company Ltd.
305 Horner Avenue
Toronto, Ontario

Société: Canadian Scale Company Limited
305 Horner Avenue
Toronto, Ontario

Manufacturer: Canadian Scale Company Ltd

Fabricant: Canadian Scale Company Ltd.

Type of Device: Railroad track scale.

Type d'appareil: Pont-bascule ferroviaire

MODEL NUMBER
No. de modèle

SECTIONAL CAPACITY
Capacité par section

CAPACITY
Capacité

MRT-428-150

120 Tons/tonnes

300 000 lbs/lb or/ou
metric equivalent/
l'équivalent métrique

MRT-7210-200

400 000 lbs/lb

Approved Accessories: This device may be used with any approved compatible accessory.

Accessoires approuvés: Le présent appareil peut être utilisé de concert avec n'importe quel accessoire compatible qui est approuvé.

Description: This device is a four section mechanical lever live deck railroad track scale.

Description: Il s'agit d'un pont-bascule ferroviaire mécanique à leviers et à tablier mobile présentant quatre sections.

The levers are "A" shaped and the weigh-bridge is suspended by a double parallel link system. Indication is by means of an existing Fairbanks-Morse Weighbeam.

Les leviers sont en forme de "A" et le châssis récepteur est soutenu par un dispositif à liaisons parallèles doubles. L'affichage du poids se fait au moyen du fléau Fairbanks-Morse en service.

Conditions of Approval:


- (1) This approval covers devices at the following locations:
Alexander Centres Ltd.,
Sudbury, Ontario
- Noranda Mines
Montreal, Quebec
- (2) No device shall be sold, leased or otherwise disposed of, for use in trade, without written disclosure to the purchaser, prior to the sale, that the device has received only temporary approval.
- (3) This approval to expire two(2) years from date of issue.
- (4) On expiry of this approval only models listed or described having the same nominal capacities and the same nominal physical dimensions and construction as an installation that has been inspected and verified by Weights and Measures, will be considered for full formal approval. After the date of expiry of the Temporary Approval if a model that has not been approved is to be installed, a new application must be submitted, with installation drawings, in order that the Temporary Approval may be reactivated and an inspection arranged.

Reference No.: G6922-C257

Conditions d'approbation:

- (1) Cet approbation couvre les appareils aux endroits suivants:
Alexander Centres Ltd.,
Sudbury, Ontario
- Noranda Mines
Montreal, Quebec
- (2) Aucun appareil ne peut être vendu, loué ou cédé de quelque autre façon sans qu'une déclaration écrite soit remise à l'acheteur, avant la vente l'informant que l'appareil n'a reçu qu'une approbation temporaire.
- (3) La présente approbation prendra fin deux(2) ans après date de son émission.
- (4) A compter de la date d'échéance de cette approbation, seuls les modèles figurant dans la liste ou décrits comme possédant la même capacité nominale, les mêmes dimensions physiques et la même construction qu'une installation inspectée et vérifiée par la Direction des poids et mesures, pourront recevoir l'approbation officielle permanente. Après la date d'échéance de la présente Approbation Temporaire, une nouvelle demande, comprenant les dessins de l'installation devra être présentée pour tout modèle, qui n'a pas été approuvé, en vue de permettre la réémission d'une approbation temporaire et de prendre les dispositions nécessaires pour l'inspection.

No. de référence: G6922-C257


APR 20 1982
W.R. Virtue

Acting Chief
Legal Metrology Branch

Chef intérimaire
Laboratoires de la métrologie légale