



Consumer and  
Corporate Affairs

Consommation et  
corporations

Legal Metrology

Métrieologie Légale

JUN  
JUN 23 1981

S.WA-X172

Approval No. d'approbation

Ottawa 4/11/80

### NOTICE OF APPROVAL – AVIS D'APPROBATION

Company Requesting Approval:

Western Scale Co. Ltd.  
1670 Kingsway Avenue  
Port Coquitlam, B.C.

Société demandant l'approbation:

Western Scale Co. Ltd.  
1670 Kingsway Avenue  
Port Coquitlam, B.C.

Type of Device: Four(4) full  
electronic grain elevator scales,  
each installation utilizing 4 load cells  
in conjunction with a digital indicator.

Appareil: Quatre(4) bascules électroniques  
d'élévateurs à grain, chaque installation  
utilisant 4 dynamomètres et un indicateur  
numérique.

Rating:

Capacité:

<u>Scale</u>	<u>Bascule Identification</u>	<u>Type</u>	<u>Capacité en tonnes Cap. tonnes</u>
C.G.C. No. 111		Receveur/Receiver	75
"	No. 112	" "	75
"	No. 113	Expéditeur/Shipper	60
"	No. 114	" "	60

Load Cells: Transducers Inc., shear  
pin type, with Western Scale model  
HSCLP-4 cross link mounting arrange-  
ment.

Dynamomètres: Transducer Inc., de  
cisaillement, à montage croisé HSCLP-4  
de Western Scale.

Indicators: Western Scale, Model DF110.

Indicateurs: Western Scale, modèle DF110.

Special Conditions: This approval is  
valid for the above-listed installations  
only, located at United Grain Growers,  
Elevator No. 3, Vancouver, B.C.


Conditions particulières: Le présent avis  
ne vise que les installations susmentionnées  
qui se trouvent à la société United Grain  
Growers, élévateur No. 3, Vancouver, B.C.

Conditions of Approval: Approval is  
granted under the Weights and Measures  
Act, S.C. 1970-71-72, Chapter 36, and  
the Weights and Measures Regulations  
(C. 1605) for use in Canada under the  
general conditions of the said Regu-  
lations and under any special condi-  
tions listed above.

Conditions d'approbation: L'approbation  
est accordée conformément à la Loi sur les  
poids et mesures, S.R.C. 1970-71-72, chapitre  
36 et au règlement d'application C.R.C.c., 1605.  
L'emploi est autorisé au Canada sous réserve  
des conditions générales dudit règlement et  
de toutes les conditions particulières  
formulées dans le présent avis.

File: G6922-W172-33

Dossier: G6922-W172-33



---

John Armstrong  
Chief

Weights and Measures Division  
Legal Metrology Branch

Chef  
Division des poids et mesures  
Direction de la Métrologie légale