



RECORD OF APPROVAL ACTION

Approval Listing Number

S.WA-T175

(Issued in advance of preparation of Formal Approval Notice)

Ottawa, January 7, 1981

Company Requestion Approval : Ramsey Rec Ltd.
Richmond Hill, Ontario

Manufacturer: Ramsey Rec.

Type of Device: Electronic Motor Truck Scale

<u>MODEL NUMBER</u>	<u>CAPACITY</u>	<u>LOAD CELL TYPE AND MODEL NUMBER</u>
10-43	200,000 lb or metric equivalent	Revere


Approved Accessories: This device may be used with any approved compatible accessory.

Device Description: This device is a full load cell truck scale with a 80'-1" x 10'-0" concrete slab deck. The load receiving element and weighbridge are supported on six(6) Revere load cells, each with a capacity of 100,000 lb. Indication is by means of a model 10-43 signal converter in a temperature controlled enclosure. Printer is a Hecc Model 522 with time/date option.

Special Conditions:

- 1) This approval covers one device only located at STELCO, Nanticoke Ontario.
- 2) No device shall be sold for use in trade without written disclosure to the purchaser, prior to the sale, that the device has received only temporary approval.
- 3) This approval expires December 31, 1981.

File No.: G6922-R218-15


John Armstrong, Chief
Weights and Measures Division
Legal Metrology Branch



S.WA-T175 Rev. 1

Approval Listing Number

Ottawa, JUL 13 1982

NOTICE OF TEMPORARY APPROVAL
Avis d'approbation temporaire

THIS SUPERCEDES RECORD OF APPROVAL ACTION
S.WA-T175 DATED January 7, 1981.

LA PRÉSENTE CIRCULAIRE REMPLACE L'AVIS
D'APPROBATION S.WA-T175 EN DATE DU 7
janvier 1981.

Company: Ramsey Rec Ltd.,
Richmond Hill, Ontario

Société: Ramsey Rec Ltd.,
Richmond Hill, Ontario

Manufacturer: Ramsey Rec Ltd.

Fabricant: Ramsey Rec Ltd.

Type of Device: Electronic motor truck
scale.

Type d'appareil: Pont-bascule routier
électronique.

MODEL NUMBER
No. de modèle

CAPACITY
Capacité

LOAD CELL(s)
Dynamomètres

10-43

200,000 lb OR/ou
Metric equivalent/
l'équivalent métrique

Revere

Approved Accessories: This device may be
used with any approved compatible
accessory.

Accessoires approuvés: Le présent
appareil peut être utilisé avec n'importe
quel accessoire approuvé et compatible.

Description: This device is a full load
cell truck scale with a 80'-1" x 10'-0"
concrete slab deck. The load receiving
element and weighbridge are supported on
six(6) Revere load cells, each with a
capacity of 100,000 lb. Indication is
by means of a model 10-43 signal
converter in a temperature controlled
enclosure. Printer is a Hecon model 522
with time/date option.

Description: Il s'agit d'un pont-bascule
routier électronique à action directe
comportant un tablier en dalles de béton
mesurant 80'-1" x 10'-0". Le tablier et
le châssis récepteur reposent sur six(6)
dynamomètres Revere qui ont chacun une
capacité de 100 000 lb. L'affichage se
fait au moyen d'un convertisseur de
signaux, modèle 10-43, installé dans un
boîtier à température contrôlée.
L'imprimante de marque Hecon, modèle 522,
assure, au choix, l'impression de l'heure
et de la date.

Conditions of Approval:

- (1) This approval covers one device only located at STELCO, Nanticoke, Ontario.
- (2) No device shall be sold, leased or otherwise disposed of, for use in trade, without written disclosure to the purchaser, prior to the sale, that the device has received only temporary approval.
- (3) This approval to expire December 31, 1982.
- (4) On expiry of this approval only models listed or described having the same nominal capacities and the same nominal physical dimensions and construction as an installation that has been inspected and verified by Weights and Measures, will be considered for full approval. After the date of expiry of this Temporary Approval if a model that has not been approved is to be installed, a new application must be submitted, with installation drawings, in order that the Temporary Approval may be reactivated and an inspection arranged.

Reference NO.: G6922-R218-15

Conditions d'approbation:

- (1) Le présent avis d'approbation ne vise que le pont-bascule routier installé à l'usine STELCO, Nanticoke, Ontario.
- (2) Aucun appareil ne doit être vendu, loué ou aliéné de toute autre façon à des fins commerciales sans que l'acheteur n'ait d'abord été avisé par écrit que l'appareil n'a reçu qu'une approbation provisoire.
- (3) Le présent avis expire le 31 décembre 1982.
- (4) Dès expiration du présent avis, seuls les modèles dont les capacités et dimensions nominales ainsi que la construction sont identiques à celles d'une installation déjà inspectée et vérifiée par la Division des poids et mesures seront considérés pour approbation officielle entière. Après la date d'expiration de l'approbation provisoire, pour qu'un modèle non approuvé soit installé, une nouvelle demande devra être présentée avec les plans de l'installation afin que l'approbation provisoire puisse être de nouveau en vigueur et qu'une date d'inspection puisse être fixée.

No. de référence: G6922-R218-15


W.R. Virtue

Acting Chief
Legal Metrology Laboratories

Chef intérimaire
Laboratoires de la Métrologie légale