



Refer to

S.WA-3078

S.WA-T177

Approval Listing Number

(Issued in advance of preparation of Formal Approval Notice)
 (Précédant l'émission de l'avis officiel d'approbation)

RECORD OF APPROVAL ACTION
Formule d'enregistrement de l'approbation

OTTAWA, 14/01/81

Company: Weightech Group
 91 Esna Drive
 Markham, Ont.

Société: Weightech Group
 91 Esna Drive
 Markham, Ont.

Manufacturer: Weightech Group

Fabricant: Weightech Group

Type of Device: Overhead Track Scale

Appareil: Pont-basculer ferroviaire aérien.

Numéro de modèle
Model Number

Capacité
Capacity

Type de dynamomètre et numéro de modèle
Load Cell Type and Model Numbers

21-5280

1500 kg x 0.5 kg

BLH model/modèle U3G1-H

Approved Accessories: Fairbanks 50-3100
 Dot Matrix Printer

Accessoires Approuvés: Imprimante par
 points Fairbanks, modèle 50-3100

Device Description: This device is a specially designed overhead track scale incorporating two (2) high temperature tension load cells. The device is used for weighing hot ingots of metal and is fitted with heat deflectors attached to the load cells. Indication is by means of a Fairbanks Digital Indicator Model H90-7302 (S.WA-917) and a Fairbanks Model 50-3100 Dot Matrix Printer.

Description: Ce pont-basculer ferroviaire aérien de conception spéciale comporte deux (2) dynamomètres de traction à température élevée. L'appareil sert à peser des lingots de métal chaud et comporte des dynamomètres équipés de déflecteurs de chaleur. L'affichage est assuré par un indicateur numérique Fairbanks, modèle H90-7302 (S.WA-917) et une imprimante par points Fairbanks, modèle 50-3100.

Conditions of Approval:


- (1) This approval covers 2 devices only those installed at Société d'Electrolyse et de Chimie Alcan Ltée, Ville de la Baie, Québec.
- (2) No device shall be sold for use in trade without written disclosure to the purchaser, prior to the sale, that the device has received only temporary approval.
- (3) This approval to expire one year from date of issue.

Reference No. *MA* G6922-F156-17

Conditions d'approbation:

- (1) Le présent avis ne vise que les deux (2) appareils installés à la Société d'Electrolyse et de Chimie Alcan Ltée, Ville de la Baie, Québec.
- (2) Aucun appareil destiné à être utilisé dans le commerce ne doit être vendu sans que l'acheteur n'ait été avisé par écrit, avant la vente, que l'appareil n'a reçu qu'une approbation provisoire.
- (3) Le présent avis expire un an après sa date de parution.

No. de référence: G6922-F156-17



John Armstrong
Chief
Weights and Measures Division
Legal Metrology Branch

Chef
Division des poids et mesures
Direction de la Métrologie légale