



Legal Metrology/Métrieologie Légale

FEB  
FEV 26 1981

RECORD OF APPROVAL ACTION

Approval Listing Number S.WA-T178

(Issued in advance of preparation of Formal Approval Notice)

OTTAWA, February 11, 1981

Company: Canadian Weigh Systems Ltd.,  
1428 Speers Road, Unit 11  
Oakville, Ontario

Manufacturer: Canadian Weigh Systems Ltd.

Type of Device: Electronic Dormant Platform Scale.

<u>MODEL NUMBER</u>	<u>SIZE</u>	<u>CAPACITY*</u>	<u>LOAD CELL</u>
866-20	6' x 6'	20 ton	H.B.H. Shear Beam
866-30	6' x 6'	30 "	
866-40	6' x 6'	40 "	
866-50	6' x 6'	50 "	50k or 20k
8610-20	6' x 10'	20 "	
8610-30	6' x 10'	30 "	
8610-40	6' x 10'	40 "	
8610-50	6' x 10'	50 "	
8614-20	6' x 14'	20 "	
8614-30	6' x 14'	30 "	
8614-40	6' x 14'	40 "	
8614-50	6' x 14'	50 "	
8620-20	6' x 20'	20 "	
8620-30	6' x 20'	30 "	
8620-40	6' x 20'	40 "	
8620-50	6' x 20'	50 "	

Approved Accessories: These devices may be used with any approved compatible accessory.

Application: The weighing of commodities in industrial trade use.


Device Description: This device is a full electronic heavy duty platform scale, utilizing H.B.H. shear beam load cells. The load cells are 20k or 50k capacity depending on the capacity of the device.

Weight indication is by means of an approved compatible electronic indicator.

Conditions of Approval:

- (1) This approval covers 10 devices only. Each device sold must be reported to the Chief, Weights and Measures by the manufacturer or distributor.
- (2) No device shall be sold for use in trade without written disclosure to the purchaser, prior to the sale, that the device has received only temporary approval.
- (3) This approval to expire one year from date of issue.

Reference Note G6922-C1209

  
\_\_\_\_\_  
John Armstrong  
Chief  
Weights and Measures Division  
Legal Metrology Branch