



Department of consumer and corporate affairs / Ministère de la consommation et des corporations



STANDARDS BRANCH - DIRECTION DES NORMES

## NOTICE OF APPROVAL

S.WA-209

Revision (1)

OTTAWA July 7, 1969.

Approved: AVERY PLATFORM FLOOR SCALE - MODEL 3205ABA

manufactured by W. & T. Avery Ltd., Birmingham 40, England.

Apparatus Listed: Portable Platform Floor Scale, with beam indication. Graduated tare bars may be fitted to the front of the beam.

Rating of Apparatus: Maximum capacity, 2700 lb. or metric equivalent.

Application: The weighing of commodities in general trade.

Description: These scales utilize two main levers of A form and cast construction. One of the main levers has an extension arm which provides a connection to the steelyard rod.

The registering beam is of steel and is notched, and is equipped with an auxiliary beam with a friction type poise. For scales requiring relatively high tare capacity, an additional graduated bar and poise may be attached to the front of the beam.

The base, platform, and beam column are constructed of cast iron. A guard frame may be fitted at the rear of the platform, in order to prevent weighed material from contacting the column.

Testing: The standard tests for platform scales shall apply.

References: SW-85-A3  
SL-102-290

Note: Approval is granted under the Weights and Measures Act, Chapter 292, and Regulations thereunder (P.C. 6894) for use in Canada under the general conditions of P.C. 6894, and under any special conditions listed above.

(for) Chief,  
Weights & Measures Division,  
Standards Branch.

R. W. MacLean,  
Director,  
Standards Branch.

