



Consumer and  
Corporate Affairs

Consommation et  
corporations

Standards

Normes

**NOTICE OF APPROVAL  
AVIS D'APPROBATION**

S.WA-210

Revision 1

Ottawa, July 26, 1976

**AVERY PORTABLE PLATFORM SCALE - MODEL 3205**

**Manufacturer:** W&T. Avery Limited  
Birmingham, England

**Apparatus:** Portable Platform Dial Scale  
Current Model - No. 3205 COJ  
Earlier Model - No. 3205 CLE

**Rating:** Maximum capacity - 2700 pounds or metric  
equivalent.

**Application:** The weighing of commodities in trade.

**Description:** These scales utilize two main levers of  
"A" form and cast construction. One of the main levers  
has an extension arm which provides a connection to the  
steelyard rod.

The indicating head is equipped with Avery  
type CLA spring resistant mechanism as approved on S.WA-387.  
The chart reading line diameter is 24 in. and 23 in. on  
Model 3205 COJ and Model 3205 CLE respectively. Scales may  
be equipped with single or dual tare beams.

The base, load platform, and dial column  
are of cast iron. The dial housing is of sheet steel  
construction. The scale may be equipped with a guard frame  
at the rear of the platform in order to prevent weighed  
material from contacting the column.

**Testing:** The standard tests for portable platform  
scales shall apply.

**Reference:** G6900-A221-23

.../2

.../2

Conditions of Approval: Approval is granted under the Weights and Measures Act, S. C. 1970-71-72, Chapter 36, and the Weights and Measures Regulations P.C. 1974-1461 of June 27, 1974 for use in Canada under the general conditions of the said Regulations, and under any special conditions listed above.

  
R. W. Maclean  
Director General  
Consumer Standards Directorate

