



S.WA-T210

Approval Listing Number

Ottawa, 1981/11/02

NOTICE OF TEMPORARY APPROVAL
Avis d'approbation temporaire

Company: Weightech Group
91 Esna Park Drive
Markham, Ontario

Société: Weightech Group
91 Esna Park Drive
Markham, Ontario

Manufacturer: Weightech Group -
Roller Conveyor by others.

Fabricant: Weightech Group -
Bascules de transporteur à
rouleaux par d'autres fabricants.

Type of Device: Motorized
conveyor scale.

Appareil: Bascule de transforma-
teur motorisée.

<u>Model No.</u>	<u>Capacity</u>	<u>Load Cells</u>
<u>No. de Modèle</u>	<u>Capacité</u>	<u>Dynamomètres</u>
SPC 3010	40 000 lb x 2 lb or/ou 18 145 kg x 1 kg	Compression

Approved Accessories: Model
50-7902 Ticket printer.

Accessoires approuvés: Impriman-
te de tickets, modèle 50-7902.

Description: This device is a
30' x 10' section of motorized
conveyor mounted on 6-20K com-
pression load cells. Weight
indication is by means of a Model
H90-7900 digital indicator
enclosed in a NEMA IV waterproof
and dustproof enclosure.

Description: Cette bascule est
un transporteur motorisé ayant une
section de 30 pi sur 10 pi et
installé sur 6 dynamomètres de
compression de 20 K. L'affichage
pondéral est assuré par un indi-
cateur numérique, modèle H90-7900
logé dans un bâti NEMA IV à
l'épreuve de l'eau et de la
poussière.

Conditions of Approval:

1. This approval only applies to
the Model SPC 3010 installed at
Air Canada, Milton, Ontario.

Conditions d'approbation:

1. La présente approbation ne
vise que le modèle SPC 3010
appartenant à Air Canada installé
à Milton en Ontario.

Conditions of Approval: (Cont'd)

2. No device shall be sold, leased or otherwise disposed of, for use in trade, without written disclosure to the purchaser, prior to the sale, that the device has received only temporary approval.

3. This approval to expire two years from date of issue.

4. On expiry of this approval only model listed or described having the same nominal capacities and the same nominal physical dimensions and construction as an installation that has been inspected and verified by Weights and Measures, will be considered for full formal approval. After the date of expiry of this Temporary Approval if a model that has not been approved is to be installed, a new application must be submitted with installation drawings, in order that the Temporary Approval may be reactivated and an inspection arranged.

Reference No.: G6922-W409

Conditions d'approbation: (Suite)

2. Aucun appareil ne peut être vendu, loué ou cédé de quelque autre façon sans qu'une déclaration écrite soit remise à l'acheteur, avant vente l'informant que l'appareil n'a reçu qu'une approbation temporaire.

3. La présente approbation prendra fin deux ans après date de son émission.

4. A compter de la date d'échéance de cette approbation, seuls les modèles figurant dans la liste ou décrits comme possédant la même capacité nominale, les mêmes dimensions physiques et la même construction qu'une installation inspectée et vérifiée par la Direction des poids et mesures, pourront recevoir l'approbation officielle permanente. Après la date d'échéance de la présente Approbation temporaire, une nouvelle demande, comprenant les dessins de l'installation devra être présentée pour tout modèle, qui n'a pas été approuvé, en vue de permettre la réémission d'une approbation temporaire et de prendre les dispositions nécessaires pour l'inspection.

Référence No.: G6922-W409



W.R. Virtue

Acting Chief
Legal Metrology Laboratories

Chef intérimaire
Laboratoires de la métrologie
légale

MAY 3 1982