

Department of consumer and corporate affairs / Ministère de la consommation et des corporations
STANDARDS BRANCH - DIRECTION DES NORMES



NOTICE OF APPROVAL

S.WA-237

Revision (1)

OTTAWA July 2, 1969.

Approved: AVERY BENCH DIAL SCALE - MODEL 3303 CFB

manufactured by W. & T. Avery Limited, Birmingham 40, England.

Apparatus Listed: Platform Bench Dial Scale, with or without beams. Dials and beams may be graduated either in avoirdupois or metric units or in both. Avoirdupois graduations may be in ounces or in decimal parts of a pound.

Rating of Apparatus: Maximum capacity, 260 lb. or metric equivalent.

Application: Weighing of commodities in general trade.

Description: These scales utilize two load levers, of "A" shape, and cast construction. The transfer lever is of similar style and is mounted below the rear main lever. Levers are borne in self-aligning suspension links.

The head mechanism may be any of the following Avery models, depending upon the vintage of the scale -

Model CFA, ref. S.WA-210

Model CLN, ref. S.WA-387

Model CIA, ref. S.WA-387.


All of the above head mechanisms utilize a single load-resisting spring, with rack and pinion indicator drive.


The weighing platform is of pressed sheet steel. The scale housing is of cast construction.

Testing: The standard tests for platform scales shall apply.

References: SW-85-A3
 SL-102-354

Note: Approval is granted under the Weights and Measures Act, Chapter 292, and Regulations thereunder (P.C. 6894) for use in Canada under the general conditions of P.C. 6894, and under any special conditions listed above.

(for) Chief, 
 Weights & Measures Division,
 Standards Branch.


 R.W. MacLean,
 Director,
 Standards Branch.

