



Consumer and
Corporate Affairs Canada

Legal Metrology

Consommation
et Corporations Canada

Métrieologie légale

APPROVAL No. - N° D'APPROBATION

S.WA-T243 Rev. 1

FEB - 4 1985

NOTICE OF TEMPORARY APPROVAL

This Notice of Temporary Approval super-
sedes Notice of Temporary Approval
S.WA-T243 dated October 19, 1982.

Company: Westcan Conveyor & Belting Ltd.
P.O. Box 593
Brandon, Ontario
R7A 5Z7

Manufacturer: Westcan Conveyor & Belting
Limited
Brandon, Ontario

Type of Device: Rotary blender scale.

MODEL NUMBER
No. de modèle

WB 5000

*or avoirdupois equivalent.

Accessories: Compatible accessories that
meet the requirements and perform in
accordance with the Weights and Measures
Act, Regulations and Specifications may
be used.

Description: This device consists of an
inclined rotating blending drum suppor-
ted at each end by a steel frame. A
shear beam load cell is positioned under
each of the four corners of the steel
frame.

Conditions of Approval:

(1) Approval covers 10 devices only.
Each device sold, leased or otherwise
disposed of, must be reported to the
Head, Mass Laboratory, Legal Metrology
Branch, Ottawa, by the Manufacturer or
Distributor.

AVIS D'APPROBATION TEMPORAIRE

Le présent avis d'approbation temporaire
remplace l'avis d'approbation temporaire
S.WA-T243 en date du 19 octobre 1982.

Société: Westcan Conveyor & Belting Ltd
P.O. Box 593
Brandon, Ontario
R7A 5Z7

Fabricant: Westcan Conveyor & Belting
Limited
Brandon, Ontario

Appareil: Bascule de malaxeur rotatif.

CAPACITY*
Capacité*

5 tons/tonnes

*ou l'équivalent en avoirdupois.

Accessoires: Tout accessoire compatible
et conforme aux exigences de construc-
tion et de rendement de la Loi, du
règlement et des prescriptions ministé-
rielles régissant les poids et mesures.

Description: Il s'agit d'un malaxeur
rotatif incliné dont les extrémités
reposent sur une structure d'acier. Un
dynamomètre de cisaillement est placé
sous chacun des quatre coins de la
structure.

Conditions d'approbation:

(1) L'approbation ne vise que 10 appa-
reils. Chaque appareil vendu, loué ou
aliéné de toute autre façon doit être
signalé au Chef du Laboratoire des
masses, Direction de la Métrieologie
légale, Ottawa, par le fabricant ou
distributeur.

- 2 -

Conditions of Approval:

(2) No device shall be sold, leased or otherwise disposed of, for use in trade, without written disclosure to the purchaser, prior to the sale, that the device has received only temporary approval.

(3) This approval to expire May 5, 1986.

(4) The blending drum must not be rotating when the product is being weighed.

Reference No.: G6922-W250

Conditions d'approbation:

(2) Aucun appareil ne doit être vendu, loué ou aliéné de toute autre façon à des fins commerciales sans que l'acheteur n'ait d'abord été avisé par écrit que l'appareil n'a reçu qu'une approbation provisoire.

(3) Le présent avis expire le 5 mai 1986.

(4) Le malaxeur ne doit pas être en rotation lorsque le produit est pesé.

No. de référence: G6922-W250


W.R. Virtue

Chief
Legal Metrology Laboratories

Chef
Laboratoires de la Métrologie légale

FEB - 4 1985