



Department of consumer and corporate affairs / Ministère de la consommation et des corporations



STANDARDS BRANCH - DIRECTION DES NORMES

NOTICE OF APPROVAL

S.WA - 256


Revision 1

OTTAWA February 28, 1972

AVERY EQUAL ARM COUNTER SCALE - MODEL 2108


- Manufacturer: W. and T. Avery Limited
Birmingham, England
- Apparatus: Equal Arm Counter Scale, Model 2108.
Scales may be equipped with single or double tare beams.
Graduations may be in avoirdupois or in metric values.
- Rating: Maximum capacity - 50 pounds or metric
equivalent.
- Application: The weighing of commodities in trade.
- Description: This scale utilizes an equal arm lever
of rectangular form and cast construction, and is equipped
with a parallel link stabilizing mechanism.
The scale base is of cast iron with enamel
finish. Tare beams are of stainless steel. The commodity
holder may be a stainless steel plate or an iron plate or
a scoop.
- Testing: The standard tests for equal arm scales
shall apply.
- References: SW - 85 - A3
SL - 102 - 368


AVERY EQUAL ARM COUNTER SCALE - MODEL 2108



- 3 -

Conditions of Approval: Approval is granted under the Weights and Measures Act, Chapter 292, and Regulations thereunder (P.C. 6894) for use in Canada under the general conditions of P.C. 6894, and under any special conditions listed above.


(for) Chief
Weights and Measures Division
Standards Branch


R. W. MacLean
Director
Standards Branch