



Department of consumer and corporate affairs / Ministère de la consommation et des corporations



STANDARDS BRANCH - DIRECTION DES NORMES

NOTICE OF APPROVAL

S.WA-261

Revision 1

OTTAWA April 8, 1970.

Approved: EVERY OVER/UNDER PLATFORM FLOOR SCALE - MODEL 3213 BDE

manufactured by W. & T. Avery Ltd., Birmingham, England.
Canadian Distributor: Robert Morse Corp. Ltd., Dominion Industrial Bldg., Ville St. Pierre, Quebec.

Apparatus Listed: Platform Floor Scale Model 3213BDE, with graduated beam and proportional weights and over and under indicator.

The indicator may be graduated plus and minus 5 pounds or plus and minus 2½ pounds.

The beam may be graduated either to 10 pounds or to 15 pounds.

Ten pound and fifteen pound beam capacities are used with 2½ pound and 5 pound indicators respectively.

Load Rating: Maximum capacity - 500 pounds.

Application: The weighing of commodities in trade.

Description: These scales are equipped with two A-shaped levers assembled in the conventional manner. Load resistance is by means of a beam and poise with counterpoise weights, and a single pendulum indicator mechanism.

Load transfer from the platform to the lever system is by means of ball and socket bearing assemblies.

The scale base and load platform are of cast iron. The beam and counterpoise weights are in a protective housing, with access by means of hinged doors at the rear of the scale.

The scale may be equipped with wheels or with levelling feet.

Testing: The standard tests for a platform floor scale shall apply.

References: SW-85-A3
SL-102-383

Note: Approval is granted under the Weights and Measures Act, Chapter 292, and Regulations thereunder (P.C. 6894) for use in Canada under the general conditions of P.C. 6894, and under any special conditions listed above.

J. H. Swanson
(for) Chief,
Weights & Measures Division,
Standards Branch.

R. W. MacLean
R. W. MacLean
Director,
Standards Branch.

