



Department of consumer and corporate affairs / Ministère de la consommation et des corporations



STANDARDS BRANCH - DIRECTION DES NORMES

NOTICE OF APPROVAL - AVIS D'APPROBATION


NO.	S.WA - 263 Revision 1
DATE	Nov. 24, 1971

AVERY OVER/UNDER BENCH SCALE - MODEL 1301 BCD

- Manufacturer: W, & T. Avery Ltd.
Birmingham, England
- Distributor: Robert Morse Corp. Ltd.
Ville St. Pierre, Quebec.
- Apparatus: Equal Arm Over/Under Scale, Model 1301 BCD.
Scales may be equipped with one or two tare beams. Beams and chart may be graduated either in avoirdupois units or metric units or in both.
- Rating: Maximum capacity - 30 pounds
- 15 kilograms
- Application: The weighing of commodities in trade.
- Description: These scales utilize a cast iron equal arm load lever with a stabilizing mechanism of the Beranger type. Load resistance is by means of a single pendulum with a cam and ribbon drive mechanism.
The scale base and indicator housing are of cast construction with enamel finish.
The commodity holder may be a platter, pan or scoop. The scale is equipped with levelling feet.
- Testing: The standard tests for equal arm counter scales shall apply.
- References: SW - 85 - A3
SL - 102 - 381

.../2


AVERY OVER/UNDER BENCH SCALE - MODEL 1301 BCD




Nov. 24, 1971

- 3 -

Conditions of Approval: Approval is granted under the
Weights and Measures Act, Chapter 292, and
Regulations thereunder (PC 6894) for use in
Canada under the general conditions of P.C.
6894, and under any special conditions
listed above.


(for) Chief,
Weights and Measures Division
Standards Branch


R. W. MacLean
Director,
Standards Branch