



Consumer and
Corporate Affairs Canada

Legal Metrology

Consommation
et Corporations Canada

Métrieologie légale

APPROVAL No. - N° D'APPROBATION

S.WA-T271 Rev. 2

JAN 18 1985

NOTICE OF TEMPORARY APPROVAL

This Notice of Temporary Approval super-
sedes Notice of Temporary Approval
S.WA-T271 revision 1 dated August 7,
1984.

Company: Howe Richardson Canada Inc.,
217 Brunswick Blvd.,
Pointe Claire, Quebec
H9R 4R7

Manufacturer: Howe Richardson Canada Inc
Pointe Claire, Quebec

Type of Device: A fully electronic motor
truck or grain dump scale that when
interfaced to an approved and compatible
digital weight indicator becomes a
weighing system.

MODEL NUMBERS

No. de modèles

PREFIX/	-3024102-	SUFFIX/
Préfixe	-3030102-	suffixe
	-3035102-	
	-5030102-	
	-5034102-	
	-5040102-	
	-5050102-	
	-6080104-	
	-8060104-	
	-8070104-	

Model Number Coding:

Prefix - "GD" = Grain Dump

"MT" = Motor Truck

1st, 2nd or 3 digit = Capacity in tons

Next 2 digits = Length in feet

Next 2 digits = Width in Feet (10 or 12)

Next digit = Number of Sections

Prefix "C" = Concrete deck

"S" = Steel deck

"W" = Wood deck

AVIS D'APPROBATION TEMPORAIRE

Le présent avis d'approbation temporaire
remplace l'avis d'approbation temporaire
S.WA-T271 Revision 1 en date du 7 août
1984.

Société: Howe Richardson Canada Inc.
217 Boul. Brunswick
Pointe Claire, Québec
H9R 4R7

Fabricant: Howe Richardson Canada Inc.
Pointe Claire, Québec

Appareil: Pont-bascule routier à trémie
à grain pour camions entièrement élec-
tronique qui, une fois relié à un indi-
cateur pondéral numérique approuvé et
compatible, devient un ensemble de
pesage.

MODEL NUMBERS

No. de modèles

PREFIX/	-8070124-	SUFFIX/
Préfixe	-8080104-	suffixe
	-8080103-	
	-8080106-	
	-8080124-	
	-10060104-	
	-10070104-	
	-10070124-	
	-10080104-	
	-10080124-	

Code du numéro de modèle:

Préfixe - "GD" = Bascule à trémie à
grain

"MT" = Pont-bascule routier

1er, 2^e ou 3^e chiffre = Capacité en
tonnes

2 chiffres suivants = Longueur en pieds

2 chiffres suivants = Largeur en pieds
(10 ou 12)

Chiffre suivant = Nombre de sections

Suffixes "C" = Tablier en béton

"S" = Tablier en acier

"W" = Tablier en bois

Canada

Accessories: Compatible accessories that meet the requirements and perform in accordance with the Weights and Measures Act, Regulations and Specifications may be used.

Description: These are fully electronic scales that use a pit installation. They are fabricated from steel, have 2, 3 or 4 sections. The platform is free floating and is restrained by use of bumper bolts. Weight is sensed using double ended shear beam load cells.

Conditions of Approval:

(1) This approval covers 10 devices only. Each device sold, leased or otherwise disposed of, must be reported to the Chief, Mass Laboratory, Legal Metrology Branch, in Ottawa, by the Manufacturer or distributor.

(2) No device shall be sold, leased or otherwise disposed of, for use in trade, without written disclosure to the purchaser, prior to the sale, that the device has received only temporary approval.

(3) This approval to expire December 31, 1985.

Reference No.: G6922-H186-15

Accessoires: Peuvent être utilisés les accessoires compatibles qui satisfont aux exigences de la Loi, du Règlement et des prescriptions ministérielles régissant les poids et mesures.

Description: Il s'agit de balances entièrement électroniques installées dans une fosse et comportant 2, 3 ou 4 sections en acier. Le tablier est de type flottant; des boulons de butée limitent ses mouvements. Des dynamomètres de cisaillement installés aux deux extrémités assurent la détection du poids.

Conditions d'approbation:

(1) La présent avis ne vise que 10 appareils. Tout appareil vendu, loué ou aliéné de toute autre façon à des fins commerciales doit être signalé au Chef du laboratoire des masses, de la Direction de la Métrologie légale, à Ottawa, par le distributeur ou le fabricant.

(2) Aucun appareil ne doit être vendu, loué ou aliéné de toute autre façon à des fins commerciales sans que l'acheteur n'ait d'abord été avisé par écrit que l'appareil n'a reçu qu'une approbation provisoire.

(3) Le présent avis expire le 31 décembre 1985.

N° de référence: G6922-H186-15


W.R. Virtue

Chief
Legal Metrology Laboratories

Chief
Laboratoire de la Métrologie légale

JAN 18 1985