



Consumer and
Corporate Affairs

Consommation et
corporations

Standards

Normes

**NOTICE OF APPROVAL
AVIS D'APPROBATION**

S.WA-289

Revision 2

Ottawa, July 22, 1976

AVERY PORTABLE PLATFORM SCALES - MODELS 3209, 4209

Manufacturer: W&T. Avery Limited,
Birmingham, England

Apparatus: Portable Platform Floor Scales -

<u>Model No.</u>	<u>Platform Size</u>
3209	36" x 36"
4209	42" x 42"

Indicating equipment is identified by means of a suffix to the basic model number.

<u>Suffix</u>	<u>Capacity Beams</u>
ABA beam optional.	Beam graduated in one unit of weight, tare
ABF	Beam of triangular section, with graduations
for 2 or 3 units of weight.	

Pillar Dials (Optional Tare Beam)

CID	Short Pillar
CFD	Short Pillar - Earlier Model
COJ	Tall Pillar
CLE	Tall Pillar - Earlier Model


Rating: Maximum capacity, 4500 pounds or metric equivalent.

Application: The weighing of commodities in trade.

.../2

SWA-289
Revision 2
July 22, 1976

AVERY PORTABLE PLATFORM SCALES - MODELS 3209. 4209



MODEL 4209 - ABA

.../3

Description: These scales are equipped with two A-shaped cast iron levers assembled in the conventional manner. Load transfer from the platform to the lever system is through ball and socket bearing assemblies.

The load platform and the lever housing are of sheet steel. Scales are mounted on wheels and equipped with a towing handle. On scales in current production the wheels are mounted inside the lever housing. Outboard wheel mounting is used on some scales of earlier manufacture.

Testing: The standard tests for platform floor scales shall apply.

Reference: G6900-A221-23

Conditions of Approval: Approval is granted under the Weights and Measures Act, S.C. 1970-71-72, Chapter 36, and the Weights and Measures Regulations P.C. 1974-1461 of June 27, 1974 for use in Canada under the general conditions of the said Regulations, and under any special conditions listed above.


R. W. MacLean
Director General
Consumer Standards Directorate

