



Department of consumer and corporate affairs / Ministère de la consommation et des corporations

STANDARDS BRANCH - DIRECTION DES NORMES

NOTICE OF APPROVAL

S.WA-290

Revision (1)

OTTAWA July 3, 1969.

Approved: AVERY PLATFORM FLOOR SCALE - MODEL 4215

manufactured by W. & T. Avery Ltd., Birmingham 40, England.

Apparatus Listed: Platform Floor Scale, Model 4215, with indicator graduated in avoirdupois or metric units. Scales may be equipped with capacity beam or with dial and unit weights. Tare beams are optional.

Rating of Apparatus: Maximum capacity, 6100 pounds or 3050 kilograms.

Application: The weighing of commodities in general trade.

Description: These scales utilize two torsion type main levers. Levers are of cast construction. Torsion shafts are hollow and of approximately square section. Load is transferred from the platform to the lever system by means of steel balls. Each corner bearing assembly utilizes two balls operating in machined sockets. The lever system is enclosed in a welded sheet steel housing.

Testing: The standard tests for platform scales shall apply.

References: SW-85-S3
SL-102-290

Note: Approval is granted under the Weights and Measures Act, Chapter 292, and Regulations thereunder (P.C. 6894) for use in Canada under the general conditions of P.C. 6894, and under any special conditions listed above.

(for) Chief,
Weights & Measures Division,
Standards Branch.

R. W. MacLean,
Director,
Standards Branch.

