



Consumer and
Corporate Affairs

Consommation et
corporations

Ottawa/Hull
K1A 0C9

Sheet 1 of 2

Consumer
Standards
Directorate

Direction
générale
des normes

RECORD OF APPROVAL ACTION

Approval listing number -

SWA-321

Addendum 1

Ottawa, October 31, 1979

Company Requesting Approval - Aurora Scale Manufacturing Ltd.
Aurora, Ontario

Manufactured by: Pennsylvania Scale Co.
21 Graybill Road,
Leola, Pennsylvania
U.S.A. 17540

Type of Device:

An equal arm metric counter scale equipped with an over and under indicator in a centre tower and a side beam.

Model No.: DC -10/50

Capacity:


See Standard chart and beam specifications on Sheet 2

The model DC 10/50 is a modernized version of the model D10. The major differences in the model DC 10/50 as compared to the model D10 are:

- i) The tower has been relocated to the centre of the scale body.
- ii) The indicator drive has been changed from a Ribbon-Cam type to a linkage-spring type.
- iii) The side beam is graduated as shown on Sheet 2.
- iv) Capacity as shown on Sheet 2.

NOTE: See Standard chart and beam specifications on Sheet 2.

File No.: G6922-P271-924

.....

 John Armstrong
 Chief,
 Weights & Measures Division
 Legal Metrology Branch

STANDARD CHART AND BEAM SPECIFICATIONS

MODEL DC-10/50

CHARTS

Over/Under, ungraduated (all capacities)

For capacities to 25 pounds

1 oz O/U x 1/8 oz

.06 lb O/U x .01 lb

1 oz O/U x .1 oz

30 g O/U x 5 g

For capacities over 25 pounds

2 oz O/U x 1/4 oz (to 40 lb)

4 oz O/U x 1/4 oz (to 50 lb)

100 g O/U x 5 g (to 22 kg)

BEAM SPECIFICATIONS

(FOR MODELS DC10/50 & DC10/50X)

1 lb x .01 lb

2 lb x 1 oz

500 g x 5 g

16 oz x 1/4 oz

8 oz x 1/8 oz

10 oz x .1 oz