

TRADE AND COMMERCE
CANADA

STANDARDS BRANCH

S-WA.417

OTTAWA, January 27, 1961

APPROVAL LISTING

THORNTON - BENCH SCALE

Under the provisions of the Weights and Measures Act, Chapter 292, R.S.C. 1952, and Regulations thereunder (P.C. 6894), the apparatus specified and illustrated herein has been listed as an approved device and may be used in Canada in accordance with the conditions applicable.


Apparatus Listed: Bench type platform scale Model 728 with a capacity of 260 pounds manufactured by F. J. Thornton and Company Limited, Birmingham, England.


Application: Weighing commodities in general trade and industrial use.

Conditions: As prescribed in P.C. 6894. The beam must be adjusted to vibrate.

Description: This device is a platform scale equipped with beam for bench use. The base and leverage system are the same as the scale listed on S-WA.192. The beam has been changed to a double bar full capacity type, one bar is graduated to 240 lbs. x 20 lbs. and the second bar to 20 lbs. x 2 ozs.

Testing: The standard tests for a platform scale shall apply.


C. S. Phillips
Chief (W. & M.)
Standards Branch


R. W. MacLean
Director
Standards Branch

Reference: A-588

(

(

THORNTON - BENCH SCALE

MODEL 738

S-WA.417

