



CANADA

3.WA-468

DEPARTMENT OF TRADE AND COMMERCE  
STANDARDS BRANCH

OTTAWA, September 23, 1963.

APPROVAL LISTING

TOLEDO SENTINEL II CYLINDER BOX TYPE SCALE

Under the provisions of the Weights and Measures Act, Chapter 292, R.S.C. 1952 and Regulations thereunder (P.C. 6894), the apparatus specified and illustrated herein has been listed as an approved device and may be used in Canada in accordance with the conditions applicable.

Apparatus Listed: Model 1361 Counter-Computing Platform Box Type Scale, manufactured by the Toledo Scale Company of Canada Limited, 2462 Howard Avenue, Windsor, Ontario.

Rating of Apparatus: Capacities - 12 pounds, 18 pounds, 24 pounds and 30 pounds.


Application: Weighing commodities in retail trade.

Conditions: As prescribed in P.C. 6894.

Description: This device is a counter-computing spring resistance box type scale. The leverage system consists of a single contour cast aluminum alloy, second class lever, supported at one end by the fulcrum bearings and suspended at the other end from the load resistance springs by a cone pivot assembly.

Testing: The standard tests for Computing Counter Scales shall apply.

  
C. S. Phillips,  
Chief, Weights & Measures Division,  
Standards Branch.

  
R. W. MacLean,  
Director,  
Standards Branch.

Ref: A-657

3

3

3

S.WA.468

TOLEDO MODEL 1361

