



Department of consumer and corporate affairs / Ministère de la consommation et des corporations



STANDARDS BRANCH - DIRECTION DES NORMES

**NOTICE OF APPROVAL
AVIS D'APPROBATION**


S.WA-490

Revision 1


OTTAWA , May 30, 1973

COUNTING ATTACHMENTS FOR DIAL SCALES


- Manufacturer : Toledo Scale,
Division of Reliance Electric Limited,
2462 Howard Ave., Windsor, Ont.
- Apparatus : Counting attachments for use on Canada Approved
Toledo Scales, models 2081, 2091, 2181, 2481, 2881
2191, 2151, 2171, 2271.
- Rating : Double Ratio Type
100:1 and 10:1 or 99:1 and 9:1
- Triple Ratio Type
1000:1, 100:1 and 10:1 or 999:1, 99:1 and 9:1
- Application : The counting of articles in trade.
- Description : These devices consist of two or three ratio
levers with load pans, which may be fitted to
the scale models as listed above.
The ratio levers are so connected that their
pan loads counterbalance loads placed on the main
platform of the scale.
Scale models 2481 and 2091 have "over-under"
dial indication and are intended for counting
service only. All other scale models listed
above may be used both for counting and weighing.
- Testing : The standard tests for counting scales shall apply.
- Reference : GL 1153-57/T170-728
- Condition of Approval : Approval is granted under the Weights and Measures
Act, R.S.C., 1970, c.W-7, and Regulations thereunder
(P.C. 6894) for use in Canada under the general
conditions of P.C. 6894, and under any special
conditions listed above.


R.W. MacLean
Director,
Standards Branch

TOLEDO COUNTING ATTACHMENTS



DOUBLE RATIO ATTACHMENT



TRIPLE RATIO ATTACHMENT