



SWA-533

Approval No. d'approbation  
Revision 1

This revision supercedes SWA-533 dated  
October 19, 1964. First approved SWA-533  
October 19, 1964.

Ottawa November 20, 1979

**NOTICE OF APPROVAL - AVIS D'APPROBATION**

TRINER FAN TYPE POSTAL SCALES

Models 803, 804, 805, 800MF, 800MG and 805PM

Company Requesting Approval:

Better Packages/Triner Scale and Manufacturing  
Company of Canada Limited  
150 Telson Road  
Markham, Ontario

Manufacturer:

Triner Scale and Manufacturing  
Company of Canada Limited  
150 Telson Road  
Markham, Ontario

Apparatus:

Postal Fan Type Counter Scales.

<u>Avoirdupois Model</u>	<u>Metric Model</u>
a. 803	d. 800MF
b. 804	e. 800MG
c. 805	f. 805PM


Rating or Capacity:

<u>Model</u>		<u>Model</u>	
803-3	3 pounds x 1/4 ounces	800MF	1 kilogram x 5 grams
804-4	4 pounds x 1/4 ounces	800MG	2 kilogram x 10 grams
805-5	5 pounds x 1 ounce	805PM	2 kilogram x 10 grams

Application:

Parcel post weighing and the determination of postal charges.

TRINER FAN TYPE POSTAL SCALES  
MODELS 803, 804 and 805



Description:


This device is a counter automatic indicating postal scale with fan shaped chart. The computations on the chart are for various postal rate classifications. The scale is of the two main lever design with an adjustable nose iron. Resistance to the load is accomplished by a double pendulum mechanism. Pivots and bearings are of hardened steel.

Testing:

The standard tests for computing counter scales shall apply.

Reference: G6922-T167-20

Conditions of Approval: As prescribed in P.C. 6894. Under the provisions of the Weights and Measures Act Chapter 292 R.S.C . 1952 and Regulations thereunder (P.C. 6894) the apparatus specified and illustrated herein have been listed as approved devices and may be used in Canada in accordance with conditions applicable.

  
-----  
John Armstrong  
Chief  
Weights and Measures Division  
Legal Metrology Branch