



CANADA

S-WA.537

DEPARTMENT OF TRADE AND COMMERCE
STANDARDS BRANCH

OTTAWA, November 2, 1964.

APPROVAL LISTING

TRINER PARCEL POST SCALE
MODEL 1825P

Under the provisions of the Weights and Measures Act, Chapter 292, R.S.C. 1952, and Regulations thereunder (P.C. 6894), the apparatus specified and illustrated herein has been listed as an approved device and may be used in Canada in accordance with the conditions applicable.

Apparatus Listed: Platform scale with fan-shaped chart model 1825P manufactured by the Triner Scale and Manufacturing Company of Canada Limited, 313 Horner Avenue, Toronto 14, Ontario.


Rating of Apparatus: Capacity 25 pounds by 1 ounce divisions.


Application: Weighing and computing postage on parcels.

Conditions: As prescribed in P.C. 6894.

Descriptions: This device is a gravity type platform scale with fan-shaped chart. The platform bears on two second class levers which are centre connected. Resistance to the load is accomplished by a double pendulum mechanism. Pivots and bearings are of hardened steel. The computations on the chart are for the various parcel post rate classifications. The scale is enclosed in a pressed steel housing. The range of the zero adjustment shall not exceed one percent of the capacity of the scale.

Testing: The standard tests for an automatic indicating scale of the fan chart type shall apply.


C. S. Phillips,
Chief, Weights and Measures Division,
Standards Branch.


R. W. MacLean,
Director,
Standards Branch.

Ref: SL-102-746.

TRINER PARCEL POST SCALE
MODEL 1825P

