



Consumer and
Corporate Affairs

Consommation et
corporations

Legal Metrology

Métrologie Légale

SWA-537

Approval No. d'approbation
Revision 1

This revision supercedes SWA-537 dated
November 2, 1964. First approved
SWA-537 dated November 2, 1964.

Ottawa November 21, 1979

NOTICE OF APPROVAL – AVIS D'APPROBATION

TRINER PARCEL POST SCALES

Model 1825P and 1800M

Company Requesting Approval:

Better Packages/Triner Scale and Manufacturing
Company of Canada Limited
150 Telson Road
Markham, Ontario

Manufacturer:

Triner Scale and Manufacturing Company
of Canada Limited
150 Telson Road
Markham, Ontario

Apparatus:

Parcel Post Scales

Avoirdupois model 1825P

Metric model 1800M

Rating or Capacity:

Model 1825P - 25 pounds x 1 ounce

Model 1800M - 10 kilograms x 20 grams

Application:


Parcel post weighing and the determination of postal charges.

Description:

This device is a gravity type platform scale with fan-shaped chart. The platform bears on two second class levers which are centre connected. Resistance to the load is accomplished by a double pendulum mechanism. Pivots and bearings are of hardened steel. The computations on the chart are for the various parcel post rate classifications. The scale is enclosed in a pressed steel housing. The range of the zero adjustment shall not exceed one percent of the capacity of the scale.

S-WA. 537

TRINER PARCEL POST SCALE
MODEL 1825P




Testing:

The standard tests for an automatic indicating scale of the fan chart type shall apply.

Reference: G6922-T167-20

Conditions of Approval: As prescribed in P.C. 6894. Under the provisions of the Weights and Measures Act, Chapter 292, R.S.C. 1952, and Regulations thereunder (P.C. 6894), the apparatus specified and illustrated herein has been listed as an approved device and may be used in Canada in accordance with the conditions applicable.



John Armstrong
Chief
Weights and Measures Division
Legal Metrology Branch