



CANADA

DEPARTMENT OF TRADE AND COMMERCE
STANDARDS BRANCH

S-WA. 541

OTTAWA, November 20, 1964.

APPROVAL LISTING

OHAUS FOUR-IN-ONE GRAIN TEST SCALE

Under the provisions of the Weights and Measures Act, Chapter 292, R.S.C. 1952, and Regulations thereunder (P.C. 6894), the apparatus specified and illustrated herein has been listed as an approved device and may be used in Canada in accordance with the conditions applicable.

Apparatus Listed: Four-In-One Grain Test Scale, Model 14A, manufactured by Ohaus Scale Corporation, 1050 Commerce Avenue, Union, New Jersey, U.S.A.


Rating of Apparatus: Maximum capacity 550 grams.


Application: Testing of grain-weight per bushel, establishing dockage and direct weighing.

Conditions: As prescribed in P.C. 6894 with the added requirement that the measuring bucket when supplied with the scale be marked in Imperial measure and bear the same serial number as the scale.

Description: This device is the same as shown on approval listing SD-WA.369, with the exception the lower percentage scale now reads 0 to 90 per cent with the zero per cent graduation aligned with the 450 gram graduation and the 90 per cent graduation aligned with the zero graduation. Formerly this scale was 0 to 100 per cent with the zero per cent graduation aligned with the 500 gram graduation and the 100 per cent graduation aligned with the zero graduation. The over and under chart assembly is now made of plastic, formerly of metal.

Testing: The standard tests for a grain test scale of this type shall apply.


C. S. Phillips,
Chief, Weights and Measures Division,
Standards Branch.


R. W. MacLean,
Director,
Standards Branch.

Ref: SL-102-191G

OHAUS FOUR-IN-ONE GRAIN TEST SCALE

